


7471A

LB87044
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41912



A= 0.0000e+000

B= 2.4101e-004

C= -1.6962e-002

Rho= 0.9998698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	104	0.000	104				
0.2	0.200	0.255	0.055	1127	0.0 %	1127				
2.5	2.500	2.459	-0.041	10274	0.0 %	10274				
5.0	5.000	4.991	-0.009	20777	0.0 %	20777				
7.5	7.500	7.403	-0.097	30788	0.0 %	30788				
10.0	10.000	10.084	0.084	41912	0.0 %	41912				

LB87044 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 25 Apr 2017 11:13:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7471A	PPB	25 Apr 2017 11:35:25	S	Std
0.2 - 1	50.2	1127	-	0.2000	7471A	PPB	25 Apr 2017 11:37:42	S	Std
2.5 - 1	52.5	10274	-	2.5000	7471A	PPB	25 Apr 2017 11:40:00	S	Std
5.0 - 1	55	20777	-	5.0000	7471A	PPB	25 Apr 2017 11:42:17	S	Std
7.5 - 1	57.5	30788	-	7.5000	7471A	PPB	25 Apr 2017 11:47:19	S	Std
10.0 - 1	510	41912	-	10.0000	7471A	PPB	25 Apr 2017 11:52:19	S	Std
ICV32 - 1	ICV32	15856	3.8045	-	7471A	PPB	25 Apr 2017 11:56:09	U	SMPL
ICB32 - 1	ICB32	291	0.0532	-	7471A	PPB	25 Apr 2017 11:58:24	U	SMPL
CCV78 - 1	CCV78	21635	5.1973	-	7471A	PPB	25 Apr 2017 12:00:40	U	SMPL
CCB78 - 1	CCB78	288	0.0524	-	7471A	PPB	25 Apr 2017 12:02:55	U	SMPL
CRA - 1	CRA	1126	0.2544	-	7471A	PPB	25 Apr 2017 12:05:11	U	SMPL
HighStd - 1	HighStd	41907	10.0831	-	7471A	PPB	25 Apr 2017 12:07:27	U	SMPL
ChkStd - 1	ChkStd	30014	7.2167	-	7471A	PPB	25 Apr 2017 12:09:43	U	SMPL
PB98447BL - 1	PBS	-33	-0.0249	-	7471A	PPB	25 Apr 2017 12:11:59	U	SMPL
PB98447BS - 1	LCSS	15660	3.7573	-	7471A	PPB	25 Apr 2017 12:14:16	U	SMPL
I2810-01 - 1	FAZTECSELECTFILL	4729	1.1228	-	7471A	PPB	25 Apr 2017 12:16:33	U	SMPL
I2811-01 - 1	FK-SF-03-001-A	4238	1.0044	-	7471A	PPB	25 Apr 2017 12:18:51	U	SMPL
I2811-03 - 1	FK-SF-03-001-B	4086	0.9678	-	7471A	PPB	25 Apr 2017 12:21:08	U	SMPL
I2811-03DUP - 1	FK-SF-03-001-BDUP	3988	0.9442	-	7471A	PPB	25 Apr 2017 12:23:26	U	SMPL
I2811-03MS - 1	FK-SF-03-001-BMS	22236	5.3421	-	7471A	PPB	25 Apr 2017 12:25:42	U	SMPL
CCV79 - 1	CCV79	21624	5.1947	-	7471A	PPB	25 Apr 2017 12:27:58	U	SMPL
CCB79 - 1	CCB79	205	0.0324	-	7471A	PPB	25 Apr 2017 12:30:14	U	SMPL
I2811-03MSD - 1	FK-SF-03-001-BMSD	20593	4.9462	-	7471A	PPB	25 Apr 2017 12:32:30	U	SMPL
I2811-05 - 1	FK-SF-03-001-C	7823	1.8685	-	7471A	PPB	25 Apr 2017 12:34:46	U	SMPL
I2813-01 - 1	E2-X10-BASE-02	366611	HIGH	-	7471A	PPB	25 Apr 2017 12:37:02	X	
I2813-02 - 1	E2-W8-BASE	3062	0.7210	-	7471A	PPB	25 Apr 2017 12:38:53	U	SMPL
I2813-03 - 1	E2-W8-SSW	16915	4.0597	-	7471A	PPB	25 Apr 2017 12:41:11	U	SMPL
I2813-04 - 1	E2-W9-SSW	1411	0.3231	-	7471A	PPB	25 Apr 2017 12:43:28	U	SMPL
I2820-07 - 1	HY-SB416	40504	9.7449	-	7471A	PPB	25 Apr 2017 12:45:45	U	SMPL
I2821-05 - 1	HY-SB426A	22657	5.4436	-	7471A	PPB	25 Apr 2017 12:48:03	U	SMPL
I2821-06 - 1	HY-SB426B	5411	1.2871	-	7471A	PPB	25 Apr 2017 12:50:21	U	SMPL
I2821-07 - 1	HY-SB427A	2150	0.5012	-	7471A	PPB	25 Apr 2017 12:52:37	U	SMPL
CCV80 - 1	CCV80	21718	5.2173	-	7471A	PPB	25 Apr 2017 12:54:53	U	SMPL
CCB80 - 1	CCB80	228	0.0380	-	7471A	PPB	25 Apr 2017 12:57:10	U	SMPL
I2821-08 - 1	HY-SB427B	8685	2.0762	-	7471A	PPB	25 Apr 2017 12:59:26	U	SMPL
PB98448BL - 1	PBS	51	-0.0047	-	7471A	PPB	25 Apr 2017 13:01:42	U	SMPL
PB98448BS - 1	LCSS	15402	3.6951	-	7471A	PPB	25 Apr 2017 13:03:59	U	SMPL
I2824-01 - 1	E2-CONCRETE-BATCH6	3485	0.8230	-	7471A	PPB	25 Apr 2017 13:06:15	U	SMPL
I2824-02 - 1	E2-CONCRETE-BATCH7	7500	1.7906	-	7471A	PPB	25 Apr 2017 13:08:32	U	SMPL
I2824-03 - 1	E2-W9-BASE	8912	2.1309	-	7471A	PPB	25 Apr 2017 13:10:49	U	SMPL
I2824-04 - 1	E2-W10-ESW	464925	HIGH	-	7471A	PPB	25 Apr 2017 13:13:06	X	
I2824-05 - 1	E2-W10-BASE	25084	6.0285	-	7471A	PPB	25 Apr 2017 13:14:58	U	SMPL
I2824-06 - 1	E2-W9-ESW	19693	4.7293	-	7471A	PPB	25 Apr 2017 13:17:17	U	SMPL
I2825-01 - 1	EME-COMPOST-2	845	0.1867	-	7471A	PPB	25 Apr 2017 13:19:34	U	SMPL
CCV81 - 1	CCV81	21583	5.1848	-	7471A	PPB	25 Apr 2017 13:21:50	U	SMPL
CCB81 - 1	CCB81	11	-0.0143	-	7471A	PPB	25 Apr 2017 13:24:06	U	SMPL
I2828-01 - 1	EME-COMPOST-02	949	0.2118	-	7471A	PPB	25 Apr 2017 13:26:23	U	SMPL
I2832-01 - 1	TEST-PIT-RD-(0-3)	1002	0.2245	-	7471A	PPB	25 Apr 2017 13:28:39	U	SMPL
I2832-01DUP - 1	TEST-PIT-RD-(0-3)DUP	968	0.2163	-	7471A	PPB	25 Apr 2017 13:30:56	U	SMPL
I2832-01MS - 1	TEST-PIT-RD-(0-3)MS	16888	4.0532	-	7471A	PPB	25 Apr 2017 13:33:12	U	SMPL
I2832-01MSD - 1	TEST-PIT-RD-(0-3)MSD	16690	4.0055	-	7471A	PPB	25 Apr 2017 13:35:29	U	SMPL
I2832-07 - 1	TEST-PIT-CURVE-(0-5)	5184	1.2324	-	7471A	PPB	25 Apr 2017 13:37:45	U	SMPL
I2835-01 - 1	MARJON-SE	566	0.1194	-	7471A	PPB	25 Apr 2017 13:40:03	U	SMPL
I2813-01DLX10 - 1		52066	12.5315	-	7471A	PPB	25 Apr 2017 13:42:21	U	SMPL
I2824-04DLX10 - 1		65251	15.7092	-	7471A	PPB	25 Apr 2017 13:44:39	U	SMPL
CCV82 - 1	CCV82	21282	5.1122	-	7471A	PPB	25 Apr 2017 13:46:56	U	SMPL
CCB82 - 1	CCB82	44	-0.0064	-	7471A	PPB	25 Apr 2017 13:49:12	U	SMPL
I2811-03LX5 - 1		880	0.1951	-	7471A	PPB	25 Apr 2017 13:51:29	U	SMPL
I2811-03A - 1		22274	5.3513	-	7471A	PPB	25 Apr 2017 13:53:45	U	SMPL
I2832-01LX5 - 1		330	0.0626	-	7471A	PPB	25 Apr 2017 13:56:02	U	SMPL
I2832-01A - 1		19164	4.6018	-	7471A	PPB	25 Apr 2017 13:58:18	U	SMPL
I2813-01DL2X50 - 1		10111	2.4199	-	7471A	PPB	25 Apr 2017 14:00:34	U	SMPL
I2824-04DL2X50 - 1		13559	3.2509	-	7471A	PPB	25 Apr 2017 14:02:51	U	SMPL
CCV83 - 1	CCV83	21274	5.1103	-	7471A	PPB	25 Apr 2017 14:05:08	U	SMPL
CCB83 - 1	CCB83	42	-0.0068	-	7471A	PPB	25 Apr 2017 14:07:25	U	SMPL