

0387045

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AK

4/25/2017 11:35

Test: Phenolics-

wg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.985	0.0	0.106	
ICB1	0.011	0.0	0.003	
CCV1	0.994	0.0	0.107	
CCB1	0.001	0.0	0.002	
PB98442BL	-0.000	0.0	0.002	
PB98442BS	1.005	0.0	0.108	
I2802-02	0.009	0.0	0.003	
I2802-02DUP	0.008	0.0	0.003	
I2802-02MS	0.828	0.0	0.089	
I2802-02MSD	0.842	0.0	0.091	
CCV2	1.019	0.0	0.110	
CCB2	-0.000	0.0	0.002	

N	12
Mean	0.475
SD	0.4947
CV%	104.11

LB 87045

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 25 11:01:20 2017

Tue Apr 25 11:30:03 2017

AL

Sample Id	Sam/Ctr/c/	Test short n:	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0537	mg/l	4/25/2017 11:01:20	
0.1PPM	A	Phenolics-	P	0.1038	mg/l	4/25/2017 11:01:21	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	4/25/2017 11:01:22	
0.50PPM	A	Phenolics-	P	0.4954	mg/l	4/25/2017 11:01:23	
1.0PPM	A	Phenolics-	P	0.989	mg/l	4/25/2017 11:01:24	
2.0PPM	A	Phenolics-	P	2.0061	mg/l	4/25/2017 11:01:25	
ICV1	S	Phenolics-	P	0.9846	mg/l	4/25/2017 11:27:15	
ICB1	S	Phenolics-	P	0.0114	mg/l	4/25/2017 11:27:16	
CCV1	S	Phenolics-	P	0.9941	mg/l	4/25/2017 11:27:17	
CCB1	S	Phenolics-	P	0.0014	mg/l	4/25/2017 11:27:18	
PB98442BL	S	Phenolics-	P	-0.0003	mg/l	4/25/2017 11:27:19	
PB98442BS	S	Phenolics-	P	1.0055	mg/l	4/25/2017 11:27:20	
I2802-02	S	Phenolics-	P	0.009	mg/l	4/25/2017 11:27:21	
I2802-02DUP	S	Phenolics-	P	0.0079	mg/l	4/25/2017 11:27:22	
I2802-02MS	S	Phenolics-	P	0.8278	mg/l	4/25/2017 11:27:23	
I2802-02MSD	S	Phenolics-	P	0.8418	mg/l	4/25/2017 11:27:24	
CCV2	S	Phenolics-	P	1.0195	mg/l	4/25/2017 11:27:25	
CCB2	S	Phenolics-	P	-0.0002	mg/l	4/25/2017 11:27:26	

LB87045

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

4/25/2017 11:03

Test Phenolics-


Accepted 4/25/2017 11:03

Factor 9.472

Bias 0.002

Coeff. of det. 0.999924

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0537	0.0500	
2	Phenol-2	0.013	0.1038	0.1000	
3	Phenol-2	0.029	0.2519	0.2500	
4	Phenol-2	0.054	0.4954	0.5000	
5	Phenol-2	0.106	0.9890	1.0000	
6	Phenol-2	0.214	2.0061	2.0000	