



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87047

Temp.: 16.5 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/25/2017	13:00
2	CAL2	1	Water			7.01	04/25/2017	13:01
3	CAL3	1	Water			10.03	04/25/2017	13:02
4	ICV	1	Water			7.00	04/25/2017	13:05
5	CCV1	1	Water			2.02	04/25/2017	13:07
6	I2847-02	1	Water			7.64	04/25/2017	13:11
7	I2847-02DUP	1	Water			7.65	04/25/2017	13:12
8	CCV2	1	Water			12.01	04/25/2017	13:13

LB 87047

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 12847 WorkList ID : 97742 Date : 4/25/2017 1:13:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/25/2017	Water	12847-02-6	Corrosivity	Cool 4 deg C	PSEG01	P62	AVENUE-A-GROUNDWATER	04/25/2017	9040C

Date/Time 04/25/17 1:13 PM
Received by: JP
Relinquished by: JP

Date/Time 04/23/17 9:00 AM
Received by: JP
Relinquished by: JP