



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87049

Temp.: 16.5 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/25/2017	13:00
2	CAL2	1	Water			7.01	04/25/2017	13:01
3	CAL3	1	Water			10.03	04/25/2017	13:02
4	ICV	1	Water			7.00	04/25/2017	13:05
5	CCV1	1	Water			2.02	04/25/2017	13:06
6	I2838-01	1	Water			3.90	04/25/2017	13:07
7	I2838-01DUP	1	Water			3.91	04/25/2017	13:08
8	I2838-02	1	Water			4.15	04/25/2017	13:10
9	I2838-03	1	Water			5.43	04/25/2017	13:11
10	CCV2	1	Water			12.01	04/25/2017	13:12

1087049

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2838

WorkList ID : 97740

Date : 4/25/2017 12:48:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2838-01	pH	Cool 4 deg C	STER09	P41	CELL-A-POOLWATER	04/24/2017	SM 4500-PH B
	Water	I2838-02	pH	Cool 4 deg C	STER09	P41	CELL-B-C-POOLWATER	04/24/2017	SM 4500-PH B
	Water	I2838-03	pH	Cool 4 deg C	STER09	P41	CELL-D-POOLWATER	04/24/2017	SM 4500-PH B

Date/Time 04/25/17 12:48 PM
 Received by: JD
 Relinquished by: JD

Date/Time 04/25/17 2:00 PM
 Received by: UP
 Relinquished by: JD