



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87067

Temp.: 19.1 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/26/2017	13:30
2	CAL2	1	Water			7.00	04/26/2017	13:31
3	CAL3	1	Water			10.02	04/26/2017	13:32
4	ICV	1	Water			6.99	04/26/2017	13:33
5	CCV1	1	Water			2.02	04/26/2017	13:34
6	I2862-01	1	Water			4.49	04/26/2017	13:35
7	I2862-01DUP	1	Water			4.50	04/26/2017	13:36
8	CCV2	1	Water			12.01	04/26/2017	13:37

LB87067

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12862

WorkList ID : 97779

Date : 4/26/2017 1:21:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12862-01	pH	Cool 4 deg C	STER04	P52	POOLWATER	04/25/2017	SM 4500-PH B

Date/Time 04/26/17 11:21:13
Received by: [Signature]
Relinquished by:

Date/Time 04/26/17 2:30 PM
Received by: [Signature]
Relinquished by: [Signature]