



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87079

Temp.: 24.1 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/26/2017	18:10
2	CAL2	1	Water			7.01	04/26/2017	18:11
3	CAL3	1	Water			10.03	04/26/2017	18:12
4	ICV	1	Water			7.01	04/26/2017	18:13
5	CCV1	1	Water			2.01	04/26/2017	18:15
6	I2870-07	1	Water			7.09	04/26/2017	18:16
7	I2870-07DUP	1	Water			7.10	04/26/2017	18:17
8	CCV2	1	Water			12.01	04/26/2017	18:19

2887079

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12870 WorkList ID : 97799 Date : 4/26/2017 6:21:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2870-07	pH	Cool 4 deg C	PSEG01	D12	PURGE-WATER-DRUMS(COI	04/26/2017	150.1

Date/Time 04/26/17 6:20 PM
Received by: [Signature]
Relinquished by: _____

Date/Time 04/26/17 7:00 PM
Received by: [Signature]
Relinquished by: [Signature]