



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87080

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/26/2017	18:10
2	CAL2	1	Water			7.01	04/26/2017	18:11
3	CAL3	1	Water			10.03	04/26/2017	18:12
4	ICV	1	Water			7.01	04/26/2017	18:13
5	CCV1	1	Water			2.02	04/26/2017	18:20
6	I2869-05	1	Solid	20.03	20	8.61	04/26/2017	18:21
7	I2869-05DUP	1	Solid	20.05	20	8.62	04/26/2017	18:22
8	I2870-02	1	Solid	20.02	20	6.92	04/26/2017	18:23
9	CCV2	1	Water			12.02	04/26/2017	18:25

287080

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2870 WorkList ID : 97800 Date : 4/26/2017 6:21:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2869-05	Corrosivity	Cool 4 deg C	PSEG01	D11	KEARNY-GEN-STA-SP-(0-2)	04/26/2017	9045C
	Solid	I2870-02	Corrosivity	Cool 4 deg C	PSEG01	D12	SOIL-CUTTINGS-DRUMS(CC	04/26/2017	9045C

Date/Time 6/26/17 6:20 PM
Received by: JP
Relinquished by: CP

Date/Time 04/26/17 7:00 PM
Received by: CP
Relinquished by: JP