



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB87087

Temp.: 23.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/27/2017	11:35
2	CAL2	1	Water			7.01	04/27/2017	11:36
3	CAL3	1	Water			10.04	04/27/2017	11:37
4	ICV	1	Water			7.01	04/27/2017	11:38
5	CCV1	1	Water			2.02	04/27/2017	11:40
6	I2860-07	1	Water			7.14	04/27/2017	11:41
7	I2860-07DUP	1	Water			7.15	04/27/2017	11:42
8	I2860-08	1	Water			6.14	04/27/2017	11:43
9	I2860-09	1	Water			7.50	04/27/2017	11:45
10	I2860-10	1	Water			7.06	04/27/2017	11:46
11	I2870-08	1	Water			7.09	04/27/2017	11:47
12	CCV2	1	Water			12.01	04/27/2017	11:48

1387087

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2860 WorkList ID : 97826 Date : 4/27/2017 11:24:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2860-07-17	Corrosivity	Cool 4 deg C	DVIR01	D21	HW-1	04/25/2017	9040C
	Water	I2860-08-17	Corrosivity	Cool 4 deg C	DVIR01	D21	HW-32	04/25/2017	9040C
	Water	I2860-09-17	Corrosivity	Cool 4 deg C	DVIR01	D21	TANK-3	04/25/2017	9040C
	Water	I2860-10-17	Corrosivity	Cool 4 deg C	DVIR01	D21	HW-3	04/25/2017	9040C
04/26/2017	Water	I2870-08-17	Corrosivity	Cool 4 deg C	PSEG01	D12	PURGE-WATER-DRUMS(COI	04/26/2017	9040C

Date/Time 04/27/17 11:24 AM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 04/27/17 11:00 AM
Received by: [Signature]
Relinquished by: [Signature]