



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87099

Temp.: 21.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/27/2017	15:25
2	CAL2	1	Water			7.00	04/27/2017	15:26
3	CAL3	1	Water			10.03	04/27/2017	15:27
4	ICV	1	Water			7.01	04/27/2017	15:28
5	CCV1	1	Water			2.02	04/27/2017	15:35
6	I2873-08	1	Water			7.08	04/27/2017	15:36
7	I2873-08DUP	1	Water			7.09	04/27/2017	15:37
8	CCV2	1	Water			12.01	04/27/2017	15:38

2587099

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i2873

WorkList ID : 97840

Date : 4/27/2017 3:15:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/27/2017	Water	I2873-08	pH	Cool 4 deg C	PSEG01	P63	FRAC-TANK-A1841DW	04/27/2017	9045C

Date/Time 04/27/17 3:15 PM
Received by: JP
Relinquished by: JP

Date/Time 04/27/17 3:15 PM
Received by: JP
Relinquished by: JP