

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87124BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54515
calibration std. phosphate 0.5 ppm	WP54514
calibration std. phosphate 0.3 ppm	WP54513
calibration std. 0 ppm	WP54510
phosphate ICV std.	WP54517
calibration std. phosphate 0.1 ppm	WP54512
calibration std. phosphate 0.05 ppm	WP54511
Combined reagent	WP54519
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate CCV std.	WP54516
Ascorbic Acid	WP54518

Intercept: -0.0005Slope: 0.6533Regression: 0.999869

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/28/2017	11:30
2	CAL2	0.05	1	50	50	0.032	04/28/2017	11:30
3	CAL3	0.10	1	50	50	0.065	04/28/2017	11:31
4	CAL4	0.30	1	50	50	0.190	04/28/2017	11:31
5	CAL5	0.50	1	50	50	0.333	04/28/2017	11:32
6	CAL6	1.00	1	50	50	0.651	04/28/2017	11:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87124

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.328	0.503	04/28/2017	11:33
2	ICB		1	50	50	0.000	0.001	04/28/2017	11:33
3	CCV1	0.50	1	50	50	0.332	0.509	04/28/2017	11:49
4	CCB1		1	50	50	0.000	0.001	04/28/2017	11:49
5	PB98553BL		1	1.01	50	0.000	0.001	04/28/2017	11:50
6	PB98553BS	0.50	1	1.02	50	0.331	0.507	04/28/2017	11:50
7	I2872-10		1	1.00	50	2.534	3.880	04/28/2017	11:51
8	I2872-10		50	1.00	50	0.515	0.789	04/28/2017	11:51
9	CCV2	0.50	1	50	50	0.334	0.512	04/28/2017	11:52
10	CCB2		1	50	50	0.000	0.001	04/28/2017	11:52