


7471A

LB 87134
INSTRUMENT ID: EX 1

Linear

 μ Abs.:

46468



A= 0.0000e+000

B= 2.1581e-004

C= 2.2775e-002

Rho= 0.9999512

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	18	0.000	18				
0.2	0.200	0.220	0.020	912	0.0 %	912				
2.5	2.500	2.488	-0.012	11422	0.0 %	11422				
5.0	5.000	4.942	-0.058	22792	0.0 %	22792				
7.5	7.500	7.473	-0.027	34524	0.0 %	34524				
10.0	10.000	10.051	0.051	46468	0.0 %	46468				

LB87134 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Apr 2017 15:39:09

Sample ID	Extended ID	μ Abs	Conc	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.00007471A	PPB	28 Apr 2017 15:40:17	S	Std
0.2 - 1	50.2	912	-	0.20007471A	PPB	28 Apr 2017 15:42:34	S	Std
2.5 - 1	52.5	11422	-	2.50007471A	PPB	28 Apr 2017 15:44:52	S	Std
5.0 - 1	55	22792	-	5.00007471A	PPB	28 Apr 2017 15:47:09	S	Std
7.5 - 1	57.5	34524	-	7.50007471A	PPB	28 Apr 2017 15:49:27	S	Std
10.0 - 1	510	46468	-	10.00007471A	PPB	28 Apr 2017 15:51:43	S	Std
ICV36 - 1	ICV36	18466	4.0079	-7471A	PPB	28 Apr 2017 15:54:25	U	SMPL
ICB36 - 1	ICB36	-66	0.0085	-7471A	PPB	28 Apr 2017 15:56:40	U	SMPL
CCV96 - 1	CCV96	22891	4.9629	-7471A	PPB	28 Apr 2017 15:58:56	U	SMPL
CCB96 - 1	CCB96	-75	0.0066	-7471A	PPB	28 Apr 2017 16:01:12	U	SMPL
CRA - 1	CRA	897	0.2164	-7471A	PPB	28 Apr 2017 16:03:28	U	SMPL
HighStd - 1	HighStd	45775	9.9015	-7471A	PPB	28 Apr 2017 16:05:43	U	SMPL
ChkStd - 1	ChkStd	31789	6.8832	-7471A	PPB	28 Apr 2017 16:08:00	U	SMPL
PB98537BL - 1	PBS	-92	0.0029	-7471A	PPB	28 Apr 2017 16:10:16	U	SMPL
PB98537BS - 1	LCSS	15213	3.3059	-7471A	PPB	28 Apr 2017 16:12:33	U	SMPL
I2869-01 - 1	KEARNY-GEN-STA-SP-(0-2)	79666	17.2155	-7471A	PPB	28 Apr 2017 16:14:50	U	SMPL
I2869-01DUP - 1	KEARNY-GEN-STA-SP-(0-2)DUP	73879	15.9666	-7471A	PPB	28 Apr 2017 16:17:07	U	SMPL
I2869-01MS - 1	KEARNY-GEN-STA-SP-(0-2)MS	93751	20.2552	-7471A	PPB	28 Apr 2017 16:19:26	U	SMPL
I2869-01MSD - 1	KEARNY-GEN-STA-SP-(0-2)MSD	95427	20.6169	-7471A	PPB	28 Apr 2017 16:24:19	U	SMPL
I2870-01 - 1	SOIL-CUTTINGS-DRUMS(COMP)	20330	4.4102	-7471A	PPB	28 Apr 2017 16:26:34	U	SMPL
CCV97 - 1	CCV97	23033	4.9935	-7471A	PPB	28 Apr 2017 16:28:51	U	SMPL
CCB97 - 1	CCB97	-39	0.0144	-7471A	PPB	28 Apr 2017 16:31:07	U	SMPL
I2873-02 - 1	MANHOLE-9-COMP	10536	2.2966	-7471A	PPB	28 Apr 2017 16:33:23	U	SMPL
I2869-01LX5 - 1		15856	3.4447	-7471A	PPB	28 Apr 2017 16:35:39	U	SMPL
I2869-01A - 1		95686	20.6728	-7471A	PPB	28 Apr 2017 16:42:15	U	SMPL
I2869-01DLX5 - 1		15820	3.4369	-7471A	PPB	28 Apr 2017 16:44:30	U	SMPL
I2869-01DUPDLX5 - 1		15019	3.2640	-7471A	PPB	28 Apr 2017 16:46:48	U	SMPL
I2869-01MSDLX5 - 1		19303	4.1886	-7471A	PPB	28 Apr 2017 16:49:05	U	SMPL
I2869-01MSDDLX5 - 1		18989	4.1208	-7471A	PPB	28 Apr 2017 16:51:22	U	SMPL
I2869-01LDLX25 - 1		3043	0.6795	-7471A	PPB	28 Apr 2017 16:53:40	U	SMPL
I2869-01ADLX5 - 1		19144	4.1542	-7471A	PPB	28 Apr 2017 16:55:58	U	SMPL
CCV98 - 1	CCV98	22899	4.9646	-7471A	PPB	28 Apr 2017 16:58:14	U	SMPL
CCB98 - 1	CCB98	-24	0.0176	-7471A	PPB	28 Apr 2017 17:00:30	U	SMPL