



## PERCENT SOLID

Analyst Name: JIGNESH  
Date: 4/29/2017

OVENTEMP IN Celsius(°C): 109  
Time IN: 17:40  
In Date: 04/28/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 104  
Time OUT: 08:40  
Out Date: 04/29/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
BalanceID: M SC-1

QC: LB87140

| Lab ID   | Client Sample ID      | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|-----------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| I2892-01 | AST-S (2.5-3)         | 1     | 1.14              | 9.53                          | 8.56                              | 88.4    |
| I2892-02 | AST-W (2-2.5)         | 2     | 1.13              | 9.97                          | 9.35                              | 93      |
| I2892-03 | AST-N (2.5-3)         | 3     | 1.18              | 9.86                          | 8.89                              | 88.8    |
| I2892-04 | AST-E (2.5-3)         | 4     | 1.13              | 9.98                          | 9.03                              | 89.3    |
| I2892-05 | OWS-W (5-5.5)         | 5     | 1.18              | 9.97                          | 8.78                              | 86.5    |
| I2892-07 | OWS-S (4.5-5)         | 6     | 1.14              | 9.91                          | 8.81                              | 87.5    |
| I2892-08 | OWS-E (4.5-5)         | 7     | 1.11              | 9.96                          | 9.04                              | 89.6    |
| I2892-09 | OWS-N (5-5.5)         | 8     | 1.16              | 9.95                          | 8.77                              | 86.6    |
| I2903-01 | 1704215273            | 9     | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-05 | SVOC-GPC-BLANK        | 10    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-06 | PEST-GPC-BLANK        | 11    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-07 | PEST-GPC-BLANK-SPIKE  | 12    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-08 | PCB-GPC-BLANK         | 13    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-09 | PCB-GPC-BLANKSPIKE    | 14    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-10 | SVOC-GPC2-BLANK       | 15    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-11 | PEST-GPC2-BLANK       | 16    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-12 | PEST-GPC2-BLANK-SPIKE | 17    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-13 | PCB-GPC2-BLANK        | 18    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2908-14 | PCB-GPC2-BLANKSPIKE   | 19    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2912-01 | NORTHWALL-042717      | 20    | 1.16              | 9.97                          | 8.33                              | 81.4    |
| I2912-02 | EASTWALL-2-042717     | 21    | 1.13              | 9.97                          | 8.64                              | 85      |
| I2914-01 | SB-03-2.5-3.5         | 22    | 1.18              | 9.97                          | 6.05                              | 55.4    |
| I2914-02 | SB-03-2.5-3.5-DUP     | 23    | 1.14              | 9.95                          | 6.06                              | 55.8    |
| I2914-03 | SB-06-COMP (0-7)      | 24    | 1.11              | 9.98                          | 3.67                              | 28.9    |
| I2914-04 | SB-06-COMP (0-7) DUP  | 25    | 1.16              | 9.54                          | 4.76                              | 43      |
| I2914-05 | SB-09-2.0-3.0         | 26    | 1.17              | 9.65                          | 4.06                              | 34.1    |
| I2914-06 | I2914-05MS            | 27    | 1.17              | 9.65                          | 4.06                              | 34.1    |
| I2914-07 | I2914-05MSD           | 28    | 1.17              | 9.65                          | 4.06                              | 34.1    |
| I2914-08 | SB-10-COMP (0-6.5)    | 29    | 1.2               | 9.84                          | 3.64                              | 28.2    |
| I2914-09 | SB-11-1.0-2.0         | 30    | 1.16              | 9.88                          | 3.49                              | 26.7    |
| I2914-10 | SB-11-COMP (0-1.0)    | 31    | 1.11              | 9.97                          | 3.84                              | 30.8    |
| I2916-01 | SB-03-COMP (0-2.5)    | 32    | 1.17              | 9.97                          | 6.77                              | 63.6    |
| I2916-02 | SB-04-7.5-8.5         | 33    | 1.17              | 9.96                          | 3.12                              | 22.2    |
| I2916-03 | SB-04-COMP (0-7.5)    | 34    | 1.16              | 9.94                          | 3.05                              | 21.5    |
| I2916-04 | SB-06-7.0-8.0         | 35    | 1.16              | 9.54                          | 2.97                              | 21.6    |
| I2916-05 | SB-09-COMP (0-2)      | 36    | 1.12              | 9.98                          | 3.77                              | 29.9    |
| I2916-06 | SB-10-6.5-7.5         | 37    | 1.17              | 9.98                          | 2.63                              | 16.6    |
| I2917-01 | SB-03-2.5-3.5         | 38    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-02 | SB-03-2.5-3.5-DUP     | 39    | 1.00              | 2.00                          | 2.00                              | 100     |



PERCENT SOLID

Analyst Name: JIGNESH  
Date: 4/29/2017

OVENTEMP IN Celsius(°C): 109  
Time IN: 17:40  
In Date: 04/28/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 104  
Time OUT: 08:40  
Out Date: 04/29/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
BalanceID: M SC-1

QC: LB87140

| Lab ID   | Client Sample ID         | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|--------------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| I2917-03 | SB-06-COMP (0-7)         | 40    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-04 | SB-06-COMP (0-7) DUP     | 41    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-05 | SB-09-2.0-3.0            | 42    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-06 | I2917-05MS               | 43    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-07 | I2917-05MSD              | 44    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-08 | SB-10-COMP (0-6.5)       | 45    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-09 | SB-11-1.0-2.0            | 46    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2917-10 | SB-11-COMP (0-1.0)       | 47    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-01 | SB-03-COMP (0-2.5)       | 48    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-02 | SB-04-7.5-8.5            | 49    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-03 | SB-04-COMP (0-7.5)       | 50    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-04 | SB-06-7.0-8.0            | 51    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-05 | SB-09-COMP (0-2)         | 52    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2918-06 | SB-10-6.5-7.5            | 53    | 1.00              | 2.00                          | 2.00                              | 100     |
| I2919-01 | TEST-PIT-COMP            | 77    | 1.17              | 9.94                          | 8.98                              | 89.1    |
| I2919-02 | TEST-PIT-GRAB            | 78    | 1.13              | 9.94                          | 8.87                              | 87.9    |
| I2919-03 | TEST-PIT-COMP-A          | 79    | 1.17              | 9.94                          | 8.98                              | 89.1    |
| I2919-04 | TEST-PIT-COMP-B          | 80    | 1.17              | 9.94                          | 8.98                              | 89.1    |
| I2919-05 | TEST-PIT-COMP            | 81    | 1.17              | 9.94                          | 8.98                              | 89.1    |
| I2920-01 | 72-11930/72-11937/VNJ-23 | 73    | 1.17              | 9.96                          | 9.17                              | 91      |
| I2920-02 | 72-11930                 | 74    | 1.1               | 9.73                          | 9.14                              | 93.2    |
| I2920-03 | 72-11930/72-11937/VNJ-23 | 75    | 1.17              | 9.96                          | 9.17                              | 91      |
| I2920-04 | 72-11930/72-11937/VNJ-23 | 76    | 1.17              | 9.96                          | 9.17                              | 91      |
| I2920-06 | VNJ-210/72-11940/72-1193 | 82    | 1.17              | 9.78                          | 9.03                              | 91.3    |
| I2920-07 | VNJ-210                  | 83    | 1.15              | 9.97                          | 9.37                              | 93.2    |
| I2920-08 | VNJ-210/72-11940/72-1193 | 84    | 1.17              | 9.78                          | 9.03                              | 91.3    |
| I2920-09 | VNJ-210/72-11940/72-1193 | 85    | 1.17              | 9.78                          | 9.03                              | 91.3    |
| I2920-11 | 144-159-284-331          | 86    | 1.11              | 9.67                          | 9.05                              | 92.8    |
| I2920-12 | 284                      | 87    | 1.16              | 9.54                          | 8.82                              | 91.4    |
| I2920-13 | 144-159-284-331          | 88    | 1.11              | 9.67                          | 9.05                              | 92.8    |
| I2920-14 | 144-159-284-331          | 89    | 1.11              | 9.67                          | 9.05                              | 92.8    |
| I2921-01 | E2-X9-BASE               | 54    | 1.16              | 9.79                          | 8.64                              | 86.7    |
| I2921-02 | E2-X10-BASE-03           | 55    | 1.19              | 9.9                           | 8.81                              | 87.5    |
| I2921-03 | E2-V8-BASE               | 56    | 1.19              | 9.91                          | 8.83                              | 87.6    |
| I2921-04 | E2-V9-BASE               | 57    | 1.17              | 9.9                           | 5.9                               | 54.2    |
| I2921-05 | E2-UST-V8N               | 58    | 1.13              | 9.98                          | 7.88                              | 76.3    |
| I2921-06 | E2-V9-ESW                | 59    | 1.13              | 9.91                          | 8.83                              | 87.7    |
| I2921-07 | E2-V10-ESW               | 60    | 1.19              | 9.79                          | 8.31                              | 82.8    |
| I2921-08 | I2921-07MS               | 61    | 1.19              | 9.79                          | 8.31                              | 82.8    |



PERCENT SOLID

Analyst Name: JIGNESH  
Date: 4/29/2017

OVENTEMP IN Celsius(°C): 109  
Time IN: 17:40  
In Date: 04/28/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 104  
Time OUT: 08:40  
Out Date: 04/29/201  
Weight Check 1.0g: 1.00 g  
Weight Check 10g: 10.00 g  
BalanceID: M SC-1

QC: LB87140

| Lab ID   | Client Sample ID          | Dish# | Dish Wt(g)<br>(A) | Dish +<br>Sample Wt(g)<br>(B) | Dish + Dry<br>Sample Wt(g)<br>(C) | % Solid |
|----------|---------------------------|-------|-------------------|-------------------------------|-----------------------------------|---------|
| I2921-09 | I2921-07MSD               | 62    | 1.19              | 9.79                          | 8.31                              | 82.8    |
| I2921-10 | E2-V10-ESW-D              | 63    | 1.11              | 9.58                          | 7.72                              | 78      |
| I2921-12 | E2-V10-BASE               | 64    | 1.18              | 9.7                           | 8.63                              | 87.4    |
| I2923-01 | 6TH-ST-SOIL-CUTTINGS (COM | 90    | 1.13              | 9.96                          | 6.63                              | 62.3    |
| I2923-02 | 6TH-ST-SOIL-CUTTINGS (GRA | 91    | 1.19              | 9.59                          | 6.62                              | 64.6    |
| I2923-03 | 6TH-ST-SOIL-CUTTINGS (GRA | 92    | 1.11              | 9.98                          | 6.96                              | 66      |
| I2923-04 | 6TH-ST-SOIL-CUTTINGS (COM | 93    | 1.13              | 9.96                          | 6.63                              | 62.3    |
| I2925-01 | 220-1W-COMP-001           | 94    | 1.14              | 9.75                          | 9.07                              | 92.1    |
| I2925-02 | 220-1W-GRAB-1             | 95    | 1.18              | 9.77                          | 9.01                              | 91.2    |
| I2925-03 | 220-1W-COMP-001-A         | 96    | 1.14              | 9.75                          | 9.07                              | 92.1    |
| I2925-04 | 220-1W-COMP-001-B         | 97    | 1.14              | 9.75                          | 9.07                              | 92.1    |
| I2925-05 | 220-1W-COMP-001           | 98    | 1.14              | 9.75                          | 9.07                              | 92.1    |
| I2925-06 | 220-1W-CONCRETE-COMP      | 99    | 1.12              | 9.64                          | 9.14                              | 94.1    |
| I2925-08 | 220-1E-CONCRETE-COMP      | 100   | 1.16              | 9.84                          | 9.38                              | 94.7    |
| I2926-03 | 280T-1                    | 65    | 1.17              | 9.95                          | 8.57                              | 84.3    |
| I2926-04 | 280T-2                    | 66    | 1.13              | 9.94                          | 9.02                              | 89.6    |
| I2927-01 | FK-GF-03-003-A            | 67    | 1.14              | 9.65                          | 8.83                              | 90.4    |
| I2927-02 | FK-GF-03-003-A            | 68    | 1.17              | 9.66                          | 8.81                              | 90      |
| I2927-03 | FK-GF-03-003-B            | 69    | 1.11              | 9.79                          | 8.92                              | 90      |
| I2927-04 | FK-GF-03-003-B            | 70    | 1.17              | 9.51                          | 8.66                              | 89.8    |
| I2928-01 | SP-1-COMP                 | 101   | 1.19              | 9.97                          | 9.00                              | 89      |
| I2928-02 | SP-1-GRAB                 | 102   | 1.15              | 9.95                          | 8.62                              | 84.9    |
| I2928-03 | SP-1-COMP                 | 103   | 1.19              | 9.97                          | 9.00                              | 89      |
| I2929-01 | GF-TEST                   | 71    | 1.12              | 9.75                          | 8.68                              | 87.6    |
| I2929-02 | GF-TEST                   | 72    | 1.17              | 9.73                          | 8.69                              | 87.9    |
| I2930-01 | RO-1-COMP                 | 104   | 1.13              | 9.74                          | 8.34                              | 83.7    |
| I2930-02 | RO-1-GRAB                 | 105   | 1.13              | 9.53                          | 8.52                              | 88      |
| I2930-03 | RO-1-COMP                 | 106   | 1.13              | 9.74                          | 8.34                              | 83.7    |
| I2930-04 | SP-1-COMP                 | 107   | 1.13              | 9.74                          | 8.55                              | 86.2    |
| I2930-05 | SP-1-GRAB                 | 108   | 1.16              | 9.7                           | 8.93                              | 91      |
| I2930-08 | SP-1-COMP                 | 109   | 1.13              | 9.74                          | 8.55                              | 86.2    |
| I2930-09 | DRUMS-COMP (1-9)          | 110   | 1.00              | 2.00                          | 2.00                              | 100     |
| I2930-12 | SP-2-COMP                 | 111   | 1.15              | 9.93                          | 8.64                              | 85.3    |
| I2930-13 | SP-2-GRAB                 | 112   | 1.19              | 9.94                          | 8.69                              | 85.7    |
| I2930-14 | SP-2-COMP                 | 113   | 1.15              | 9.93                          | 8.64                              | 85.3    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

LB 87140

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample       | Collect Date | Method       |
|----------|--------|----------|----------------|--------------|----------|------------------|-----------------------|--------------|--------------|
|          | Solid  | I2892-01 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | AST-S(2.5-3)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-02 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | AST-W(2-2.5)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-03 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | AST-N(2.5-3)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-04 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | AST-E(2.5-3)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-05 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | OWS-W(5-5.5)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-07 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | OWS-S(4.5-5)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-08 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | OWS-E(4.5-5)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2892-09 | Percent Solids | Cool 4 deg C | WOOD06   | D31              | OWS-N(5-5.5)          | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2903-01 | Percent Solids | Cool 4 deg C | PHAR01   | D11              | 1704215273            | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2908-05 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | SVOC-GPC-BLANK        | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-06 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PEST-GPC-BLANK        | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-07 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PEST-GPC-BLANK-SPIKE  | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-08 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PCB-GPC-BLANK         | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-09 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PCB-GPC-BLANK-SPIKE   | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-10 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | SVOC-GPC2-BLANK       | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-11 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PEST-GPC2-BLANK       | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-12 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PEST-GPC2-BLANK-SPIKE | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-13 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PCB-GPC2-BLANK        | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2908-14 | Percent Solids | Cool 4 deg C | CHEM02   | D31              | PCB-GPC2-BLANK-SPIKE  | 04/21/2017   | Chemtech -SO |
|          | Solid  | I2912-01 | Percent Solids | Cool 4 deg C | OBR105   | D41              | NORTHWALL-042717      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2912-02 | Percent Solids | Cool 4 deg C | OBR105   | D41              | EASTWALL-2-042717     | 04/27/2017   | Chemtech -SO |

Date/Time 04/28/17 4:00 PM

Received by: JP

Relinquished by: JP

Date/Time 04/28/17 5:35 PM

Received by: JP

Relinquished by: JP

# WORKLIST(Hardcopy Internal Chain)

2587140

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample    | Collect Date | Method       |
|----------|--------|----------|----------------|--------------|----------|------------------|--------------------|--------------|--------------|
|          | Solid  | I2914-01 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-03-2.5-3.5      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-02 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-03-2.5-3.5-DUP  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-03 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-06-COMP(0-7)    | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-04 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-06-COMP(0-7)DUP | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-05 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-09-2.0-3.0      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-06 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | I2914-05MS         | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-07 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | I2914-05MSD        | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-08 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-10-COMP(0-6.5)  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-09 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-11-1.0-2.0      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2914-10 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-11-COMP(0-1.0)  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-01 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-03-COMP(0-2.5)  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-02 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-04-7.5-8.5      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-03 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-04-COMP(0-7.5)  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-04 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-06-7.0-8.0      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-05 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-09-COMP(0-2)    | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2916-06 | Percent Solids | Cool 4 deg C | LIRO01   | D32              | SB-10-6.5-7.5      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2917-01 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-03-2.5-3.5      | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2917-02 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-03-2.5-3.5-DUP  | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2917-03 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-06-COMP(0-7)    | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2917-04 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-06-COMP(0-7)DUP | 04/27/2017   | Chemtech -SO |
|          | Solid  | I2917-05 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-09-2.0-3.0      | 04/27/2017   | Chemtech -SO |

Date/Time 04/28/17 11:00 AM

Received by: JP

Relinquished by: CP

Date/Time 04/28/17 3:35 PM

Received by: CP

Relinquished by: JP

2387140

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample           | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|---------------------------|--------------|--------------|
|            | Solid  | I2917-06 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | I2917-05MS                | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2917-07 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | I2917-05MSD               | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2917-08 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-10-COMP(0-6.5)         | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2917-09 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-11-1.0-2.0             | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2917-10 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-11-COMP(0-1.0)         | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-01 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-03-COMP(0-2.5)         | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-02 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-04-7.5-8.5             | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-03 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-04-COMP(0-7.5)         | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-04 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-06-7.0-8.0             | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-05 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-09-COMP(0-2)           | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2918-06 | Percent Solids | Cool 4 deg C | LIRO01   | A11              | SB-10-6.5-7.5             | 04/27/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2919-01 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | TEST-PIT-COMP             | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2919-02 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | TEST-PIT-GRAB             | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2919-03 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | TEST-PIT-COMP-A           | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2919-04 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | TEST-PIT-COMP-B           | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2919-05 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | TEST-PIT-COMP             | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-01 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 72-11930/72-11937/VNJ-239 | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-02 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 72-11930                  | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-03 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 72-11930/72-11937/VNJ-239 | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-04 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 72-11930/72-11937/VNJ-239 | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-06 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | VNJ-210/72-11940/72-11932 | 04/28/2017   | Chemtech -SO |

Date/Time 04/28/17 4:00 PM

Received by: JD

Relinquished by: GJ

Date/Time 04/28/17 5:35 PM

Received by: CD

Relinquished by: JD

# WORKLIST(Hardcopy Internal Chain)

1087146

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample           | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|---------------------------|--------------|--------------|
| 05/03/2017 | Solid  | I2920-07 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | VNJ-210                   | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-08 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | VNJ-210/72-11940/72-11932 | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-09 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | VNJ-210/72-11940/72-11932 | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-11 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 144-159-284-331           | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-12 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 284                       | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-13 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 144-159-284-331           | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2920-14 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 144-159-284-331           | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-01 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-X9-BASE                | 04/26/2017   | Chemtech -SO |
|            | Solid  | I2921-02 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-X10-BASE-03            | 04/26/2017   | Chemtech -SO |
|            | Solid  | I2921-03 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V8-BASE                | 04/26/2017   | Chemtech -SO |
|            | Solid  | I2921-04 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V9-BASE                | 04/26/2017   | Chemtech -SO |
| 05/07/2017 | Solid  | I2921-05 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-UST-V8N                | 04/27/2017   | Chemtech -SO |
|            | Solid  | I2921-06 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V9-ESW                 | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-07 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V10-ESW                | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-08 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | I2921-07MS                | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-09 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | I2921-07MSD               | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-10 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V10-ESW-D              | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2921-12 | Percent Solids | Cool 4 deg C | E2PR01   | D31              | E2-V10-BASE               | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2923-01 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 6TH-ST-SOIL-CUTTINGS(CO   | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2923-02 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 6TH-ST-SOIL-CUTTINGS(GR   | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2923-03 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 6TH-ST-SOIL-CUTTINGS(GR   | 04/28/2017   | Chemtech -SO |

Date/Time 04/28/17 11:00 AM

Received by: JP

Relinquished by: UA

Date/Time 04/28/17 5:38 PM

Received by: CD

Relinquished by: JP

213 84140

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample         | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-------------------------|--------------|--------------|
| 05/03/2017 | Solid  | I2923-04 | Percent Solids | Cool 4 deg C | PSEG01   | P62              | 6TH-ST-SOIL-CUTTINGS(CO | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-01 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-COMP-001         | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-02 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-GRAB-1           | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-03 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-COMP-001-A       | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-04 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-COMP-001-B       | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-05 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-COMP-001         | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-06 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1W-CONCRETE-COMP    | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2925-08 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | 220-1E-CONCRETE-COMP    | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2926-03 | Percent Solids | Cool 4 deg C | DIST01   | D42              | 280T-1                  | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2926-04 | Percent Solids | Cool 4 deg C | DIST01   | D42              | 280T-2                  | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2927-01 | Percent Solids | Cool 4 deg C | TULL02   | D32              | FK-GF-03-003-A          | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2927-02 | Percent Solids | Cool 4 deg C | TULL02   | D32              | FK-GF-03-003-A          | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2927-03 | Percent Solids | Cool 4 deg C | TULL02   | D32              | FK-GF-03-003-B          | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2927-04 | Percent Solids | Cool 4 deg C | TULL02   | D32              | FK-GF-03-003-B          | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2928-01 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | SP-1-COMP               | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2928-02 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | SP-1-GRAB               | 04/28/2017   | Chemtech -SO |
| 05/03/2017 | Solid  | I2928-03 | Percent Solids | Cool 4 deg C | PSEG01   | P63              | SP-1-COMP               | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2929-01 | Percent Solids | Cool 4 deg C | TULL02   | D31              | GF-TEST                 | 04/28/2017   | Chemtech -SO |
|            | Solid  | I2929-02 | Percent Solids | Cool 4 deg C | TULL02   | D31              | GF-TEST                 | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-01 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | RO-1-COMP               | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-02 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | RO-1-GRAB               | 04/28/2017   | Chemtech -SO |

Date/Time 04/28/17 4:00PM Date/Time 04/28/17 3:55PM  
 Received by: JP Received by: CP  
 Relinquished by: CP Relinquished by: JP

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042817

WorkList ID : 97854

Date : 4/28/2017 8:28:02 AM

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|-----------------|--------------|--------------|
| 05/05/2017 | Solid  | I2930-03 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | RO-1-COMP       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-04 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-1-COMP       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-05 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-1-GRAB       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-08 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-1-COMP       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-09 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | DRUMS-COMP(1-9) | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-12 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-2-COMP       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-13 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-2-GRAB       | 04/28/2017   | Chemtech -SO |
| 05/05/2017 | Solid  | I2930-14 | Percent Solids | Cool 4 deg C | PSEG01   | P61              | SP-2-COMP       | 04/28/2017   | Chemtech -SO |

Date/Time 04/28/17 11:00 AM

Received by: JP

Relinquished by: CP

Date/Time 04/28/17 5:35 PM

Received by: CP

Relinquished by: JP