



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87161

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	04/29/2017	12:00
2	CAL2	1	Water			7.01	04/29/2017	12:01
3	CAL3	1	Water			10.03	04/29/2017	12:02
4	ICV	1	Water			7.00	04/29/2017	12:03
5	CCV1	1	Water			2.01	04/29/2017	12:30
6	I2919-05	1	Solid	20.02	20	9.26	04/29/2017	12:31
7	I2919-05DUP	1	Solid	20.04	20	9.27	04/29/2017	12:32
8	I2920-04	1	Solid	20.01	20	8.71	04/29/2017	12:35
9	I2920-09	1	Solid	20.04	20	9.38	04/29/2017	12:36
10	I2920-14	1	Solid	20.01	20	8.66	04/29/2017	12:37
11	I2923-04	1	Solid	20.03	20	8.71	04/29/2017	12:39
12	I2925-05	1	Solid	20.05	20	8.20	04/29/2017	12:41
13	I2928-03	1	Solid	20.04	20	6.94	04/29/2017	12:42
14	I2930-03	1	Solid	20.02	20	6.80	04/29/2017	12:43
15	I2930-08	1	Solid	20.04	20	7.64	04/29/2017	12:45
16	CCV2	1	Water			12.02	04/29/2017	12:46
17	I2930-10	1	Solid	20.01	20	7.35	04/29/2017	12:48
18	I2930-14	1	Solid	20.04	20	8.14	04/29/2017	12:50
19	CCV3	1	Water			12.01	04/29/2017	12:51

2B87(6)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2919

WorkList ID : 97913

Date : 4/29/2017 9:02:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/03/2017	Solid	I2919-05	Corrosivity	Cool 4 deg C	PSEG01	P62	TEST-PIT-COMP	04/28/2017	9045C
05/03/2017	Solid	I2920-04	Corrosivity	Cool 4 deg C	PSEG01	P62	72-11930/72-11937/VNJ-239	04/28/2017	9045C
05/03/2017	Solid	I2920-09	Corrosivity	Cool 4 deg C	PSEG01	P62	VNJ-210/72-11940/72-11932	04/28/2017	9045C
05/03/2017	Solid	I2920-14	Corrosivity	Cool 4 deg C	PSEG01	P62	144-159-284-331	04/28/2017	9045C
05/03/2017	Solid	I2923-04	Corrosivity	Cool 4 deg C	PSEG01	P62	6TH-ST-SOIL-CUTTINGS(CO	04/28/2017	9045C
05/03/2017	Solid	I2925-05	Corrosivity	Cool 4 deg C	PSEG01	P63	220-1W-COMP-001	04/28/2017	9045C
05/03/2017	Solid	I2928-03	Corrosivity	Cool 4 deg C	PSEG01	P63	SP-1-COMP	04/28/2017	9045C
05/05/2017	Solid	I2930-03	Corrosivity	Cool 4 deg C	PSEG01	P61	RO-1-COMP	04/28/2017	9045C
05/05/2017	Solid	I2930-08	Corrosivity	Cool 4 deg C	PSEG01	P61	SP-1-COMP	04/28/2017	9045C
05/05/2017	Solid	I2930-10	Corrosivity	Cool 4 deg C	PSEG01	P61	DRUMS-COMP(1-9)	04/28/2017	9045C
05/05/2017	Solid	I2930-14	Corrosivity	Cool 4 deg C	PSEG01	P61	SP-2-COMP	04/28/2017	9045C

Date/Time 6/24/17 9:00 PM
 Received by: JH
 Relinquished by: [Signature]

Date/Time 6/24/17 9:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]