



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87163

Temp.: 19.4 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/29/2017	09:25
2	CAL2	1	Water			7.01	04/29/2017	09:26
3	CAL3	1	Water			10.02	04/29/2017	09:27
4	ICV	1	Water			6.99	04/29/2017	09:30
5	CCV1	1	Water			2.02	04/29/2017	09:31
6	I2923-06	1	Water			6.87	04/29/2017	09:32
7	I2923-06DUP	1	Water			6.88	04/29/2017	09:33
8	CCV2	1	Water			12.01	04/29/2017	09:34

2887163

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12923

WorkList ID : 97916

Date : 4/29/2017 9:04:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/28/2017	Water	I2923-06	pH	Cool 4 deg C	PSEG01	P62	6TH-ST-PURGE-WATER(CON	04/28/2017	9045C

Date/Time 04/29/17 9:05 AM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 04/29/17 11:00 AM
Received by: [Signature]
Relinquished by: [Signature]