

LB87171


7470A

INSTRUMENT ID : CV 1

Linear

 μ Abs.:

34978



A= 0.0000e+000

B= 2.8394e-004

C= -5.1267e-002

Rho= 0.9997373

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	184	0.000	184				
0.2	0.200	0.193	-0.007	860	0.0 %	860				
2.5	2.500	2.422	-0.078	8711	0.0 %	8711				
5.0	5.000	5.071	0.071	18040	0.0 %	18040				
7.5	7.500	7.632	0.132	27060	0.0 %	27060				
10.0	10.000	9.881	-0.119	34978	0.0 %	34978				

LB87171 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 May 2017 13:58:04

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SP	184	-	0.00007470A	PPB	01 May 2017 14:07:41	S	Std
0.2 - 1	SP-2	860	-	0.20007470A	PPB	01 May 2017 14:09:49	S	Std
2.5 - 1	SP-5	8711	-	2.50007470A	PPB	01 May 2017 14:11:57	S	Std
5.0 - 1	SP-5	18040	-	5.00007470A	PPB	01 May 2017 14:14:06	S	Std
7.5 - 1	SP-5	27060	-	7.50007470A	PPB	01 May 2017 14:16:13	S	Std
10.0 - 1	SP-10	34978	-	10.00007470A	PPB	01 May 2017 14:20:45	S	Std
ICV37 - 1	ICV37	13751	3.8533	-7470A	PPB	01 May 2017 14:23:16	U	SMPL
ICB37 - 1	ICB37	41	-0.0396	-7470A	PPB	01 May 2017 14:25:22	U	SMPL
CCV01 - 1	CCV01	18089	5.0850	-7470A	PPB	01 May 2017 14:27:28	U	SMPL
CCB01 - 1	CCB01	107	-0.0209	-7470A	PPB	01 May 2017 14:29:35	U	SMPL
CRA - 1	CRA	873	0.1966	-7470A	PPB	01 May 2017 14:31:41	U	SMPL
HighStd - 1	HighStd	35412	10.0038	-7470A	PPB	01 May 2017 14:33:48	U	SMPL
ChkStd - 1	ChkStd	24931	7.0278	-7470A	PPB	01 May 2017 14:35:55	U	SMPL
PB98564BL - 1	PBW	146	-0.0098	-7470A	PPB	01 May 2017 14:38:03	U	SMPL
PB98564BS - 1	LCSW	13431	3.7624	-7470A	PPB	01 May 2017 14:40:10	U	SMPL
PB98564TB - 1	PB98534TB	113	-0.0192	-7470A	PPB	01 May 2017 14:42:18	U	SMPL
I2873-07 - 1	MANHOLE-9-COMP	148	-0.0092	-7470A	PPB	01 May 2017 14:44:26	U	SMPL
I2873-07DUP - 1	MANHOLE-9-COMPDUP	163	-0.0050	-7470A	PPB	01 May 2017 14:46:34	U	SMPL
I2873-07MS - 1	MANHOLE-9-COMPMS	13676	3.8320	-7470A	PPB	01 May 2017 14:48:43	U	SMPL
I2873-07MSD - 1	MANHOLE-9-COMPMSD	13286	3.7212	-7470A	PPB	01 May 2017 14:50:50	U	SMPL
CCV02 - 1	CCV02	18310	5.1478	-7470A	PPB	01 May 2017 14:52:57	U	SMPL
CCB02 - 1	CCB02	100	-0.0229	-7470A	PPB	01 May 2017 14:55:04	U	SMPL
I2873-09 - 1	FRAC-TANK-A1841DW	28	-0.0433	-7470A	PPB	01 May 2017 14:57:10	U	SMPL
I2895-15 - 1	SB-01-COMP(0-2.0)	108	-0.0206	-7470A	PPB	01 May 2017 14:59:17	U	SMPL
I2895-16 - 1	SB-02-COMP(0-3.0)	180	-0.0002	-7470A	PPB	01 May 2017 15:01:23	U	SMPL
I2895-17 - 1	SB-05-COMP(0-4.0)	635	0.1290	-7470A	PPB	01 May 2017 15:03:30	U	SMPL
I2895-18 - 1	SB-05-COMP(4.0-6.0)	324	0.0407	-7470A	PPB	01 May 2017 15:05:38	U	SMPL
I2895-19 - 1	SB-07-COMP(0-5.5)	264	0.0237	-7470A	PPB	01 May 2017 15:07:46	U	SMPL
I2895-20 - 1	SB-08-COMP(0-2.0)	229	0.0138	-7470A	PPB	01 May 2017 15:09:54	U	SMPL
I2895-21 - 1	SB-12-COMP(0-2.0)	217	0.0103	-7470A	PPB	01 May 2017 15:12:02	U	SMPL
I2873-07LX5 - 1		168	-0.0036	-7470A	PPB	01 May 2017 15:14:11	U	SMPL
I2873-07A - 1		13498	3.7814	-7470A	PPB	01 May 2017 15:16:18	U	SMPL
CCV03 - 1	CCV03	17839	5.0140	-7470A	PPB	01 May 2017 15:18:24	U	SMPL
CCB03 - 1	CCB03	20	-0.0456	-7470A	PPB	01 May 2017 15:20:31	U	SMPL