



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB87174

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/01/2017	16:10
2	CAL2	1	Water			7.00	05/01/2017	16:11
3	CAL3	1	Water			10.02	05/01/2017	16:12
4	ICV	1	Water			7.01	05/01/2017	16:15
5	CCV1	1	Water			2.02	05/01/2017	16:16
6	I2936-06	1	Solid	20.02	20	9.17	05/01/2017	16:17
7	I2936-06DUP	1	Solid	20.03	20	9.18	05/01/2017	16:18
8	I2936-12	1	Solid	20.03	20	9.11	05/01/2017	16:20
9	I2936-18	1	Solid	20.02	20	9.03	05/01/2017	16:21
10	I2940-03	1	Solid	20.04	20	7.81	05/01/2017	16:22
11	CCV2	1	Water			12.01	05/01/2017	16:23

LB 87174

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2936 WorkList ID : 97990 Date : 5/1/2017 3:31:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2936-06	Corrosivity	Cool 4 deg C	DEWB01	D51	TP2-COMP(1-10FT)	05/01/2017	9045C
	Solid	I2936-12	Corrosivity	Cool 4 deg C	DEWB01	D51	TP3-COMP(0.5-4.5FT)	05/01/2017	9045C
	Solid	I2936-18	Corrosivity	Cool 4 deg C	DEWB01	D51	TP6-COMP(0.5-5FT)	05/01/2017	9045C
05/04/2017	Solid	I2940-03	Corrosivity	Cool 4 deg C	PSEG01	P12	5040-COMP	05/01/2017	9045C

Date/Time 05/01/17 3:31/17
Received by: JP
Relinquished by:

Date/Time 05/01/17 7:00 D
Received by: JP
Relinquished by: JP