



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87188

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/26/2017	16:00
2	CAL2	1	Water			7.01	04/26/2017	16:01
3	CAL3	1	Water			10.03	04/26/2017	16:02
4	ICV	1	Water			7.00	04/26/2017	16:05
5	CCV1	1	Water			2.01	04/26/2017	16:06
6	I2855-01	1	Solid	20.03	20	7.17	04/26/2017	16:07
7	I2855-01DUP	1	Solid	20.04	20	7.18	04/26/2017	16:09
8	I2855-02	1	Solid	20.01	20	5.97	04/26/2017	16:10
9	CCV2	1	Water			12.01	04/26/2017	16:11

2887188

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2855 WorkList ID : 97793 Date : 4/26/2017 3:36:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2855-01	Corrosivity	Cool 4 deg C	CLEA11	D22	OWS-2	04/25/2017	9045C
	Solid	I2855-02	Corrosivity	Cool 4 deg C	CLEA11	D22	OWS-1	04/25/2017	9045C

Date/Time 04/26/17 5:35PM
 Received by: CR
 Relinquished by: JP

Date/Time 04/26/17 3:36PM
 Received by: JP
 Relinquished by: CR