


ISM02.4

LB87208
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47449



A= 0.0000e+000

B= 2.1155e-004

C= 1.4138e-002

Rho= 0.9999491

Accept=Accepted

slope
y=1.015e3x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	0.030	0.030	74	0.000	74					6
0.2	0.200	0.213	0.013	938	0.0 %	938					0
2.5	2.500	2.495	-0.005	11727	0.0 %	11727					-1
5.0	5.000	4.939	-0.061	23281	0.0 %	23281					0
7.5	7.500	7.471	-0.029	35248	0.0 %	35248					0
10.0	10.000	10.052	0.052	47449	0.0 %	47449					1

NB
5/3/17

LB87208
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	74	-		0 ISM02.4	PPB	5/3/2017 8:21	Std
	0.2 S0.2	938	-		0.2 ISM02.4	PPB	5/3/2017 8:23	Std
	2.5 S2.5	11727	-		2.5 ISM02.4	PPB	5/3/2017 8:26	Std
	5 S5	23281	-		5 ISM02.4	PPB	5/3/2017 8:28	Std
	7.5 S7.5	35248	-		7.5 ISM02.4	PPB	5/3/2017 8:30	Std
	10 S10	47449	-		10 ISM02.4	PPB	5/3/2017 8:32	Std
ICV	ICV	18376	3.9017 -		ISM02.4	PPB	5/3/2017 8:35	SMPL
ICB	ICB	5	0.0152 -		ISM02.4	PPB	5/3/2017 8:38	SMPL
CCV023	CCV023	23623	5.0117 -		ISM02.4	PPB	5/3/2017 8:40	SMPL
CCB023	CCB023	13	0.0169 -		ISM02.4	PPB	5/3/2017 8:42	SMPL
PB98659BL	PBS004	23	0.019 -		ISM02.4	PPB	5/3/2017 8:44	SMPL
I2876-01	MF4C16	547	0.1299 -		ISM02.4	PPB	5/3/2017 8:47	SMPL
I2876-02	MF4C16D	435	0.1062 -		ISM02.4	PPB	5/3/2017 8:49	SMPL
I2876-03	MF4C16S	12163	2.5873 -		ISM02.4	PPB	5/3/2017 8:51	SMPL
I2876-04	MF4C17	2294	0.4994 -		ISM02.4	PPB	5/3/2017 8:53	SMPL
I2876-05	MF4C18	230	0.0628 -		ISM02.4	PPB	5/3/2017 8:56	SMPL
I2876-06	MF4C19	257	0.0685 -		ISM02.4	PPB	5/3/2017 8:58	SMPL
I2876-07	MF4C20	169	0.0499 -		ISM02.4	PPB	5/3/2017 9:00	SMPL
I2876-08	MF4C21	539	0.1282 -		ISM02.4	PPB	5/3/2017 9:03	SMPL
I2876-09	MF4C22	216	0.0598 -		ISM02.4	PPB	5/3/2017 9:05	SMPL
I2876-10	MF4C23	251	0.0672 -		ISM02.4	PPB	5/3/2017 9:07	SMPL
I2876-11	MF4C24	197	0.0558 -		ISM02.4	PPB	5/3/2017 9:09	SMPL
I2876-12	MF4C25	148	0.0454 -		ISM02.4	PPB	5/3/2017 9:12	SMPL
I2876-13	MF4C10	503	0.1205 -		ISM02.4	PPB	5/3/2017 9:14	SMPL
I2876-14	MF4C11	139	0.0435 -		ISM02.4	PPB	5/3/2017 9:16	SMPL
I2876-15	MF4C12	157	0.0474 -		ISM02.4	PPB	5/3/2017 9:18	SMPL
I2876-16	MF4C13	135	0.0427 -		ISM02.4	PPB	5/3/2017 9:21	SMPL
I2876-17	MF4C14	180	0.0522 -		ISM02.4	PPB	5/3/2017 9:23	SMPL
I2876-18	MF4E27	126	0.0408 -		ISM02.4	PPB	5/3/2017 9:25	SMPL
I2876-19	MF4E28	110	0.0374 -		ISM02.4	PPB	5/3/2017 9:28	SMPL
CCV024	CCV024	24114	5.1156 -		ISM02.4	PPB	5/3/2017 9:30	SMPL
CCB024	CCB024	51	0.0249 -		ISM02.4	PPB	5/3/2017 9:32	SMPL
I2876-20	MF4E29	266	0.0704 -		ISM02.4	PPB	5/3/2017 9:34	SMPL
I2876-21	MF4E30	64	0.0277 -		ISM02.4	PPB	5/3/2017 9:37	SMPL
I2876-22	MF4E31	1469	0.3249 -		ISM02.4	PPB	5/3/2017 9:39	SMPL
PB98660BL	PBS004	7	0.0156 -		ISM02.4	PPB	5/3/2017 9:41	SMPL
I2910-02	MF4E19	354	0.089 -		ISM02.4	PPB	5/3/2017 9:44	SMPL
I2910-03	MF4E19D	343	0.0867 -		ISM02.4	PPB	5/3/2017 9:46	SMPL
I2910-04	MF4E19S	12151	2.5847 -		ISM02.4	PPB	5/3/2017 9:48	SMPL
I2910-05	MF4E20	201	0.0567 -		ISM02.4	PPB	5/3/2017 9:50	SMPL
I2910-06	MF4E21	223	0.0613 -		ISM02.4	PPB	5/3/2017 9:53	SMPL
I2910-07	MF4E22	220	0.0607 -		ISM02.4	PPB	5/3/2017 9:55	SMPL
I2910-08	MF4E23	397	0.0981 -		ISM02.4	PPB	5/3/2017 9:57	SMPL
I2910-09	MF4E24	318	0.0814 -		ISM02.4	PPB	5/3/2017 10:00	SMPL
I2910-10	MF4E25	21694	4.6036 -		ISM02.4	PPB	5/3/2017 10:02	SMPL
I2910-11	MF4E26	44743	9.4797 -		ISM02.4	PPB	5/3/2017 10:04	SMPL

LB87208
INSTRUMENT ID : CV1

I2910-12	MF4E43	691	0.1603 -	ISM02.4	PPB	5/3/2017 10:06 SMPL
I2910-13	MF4E44	314	0.0806 -	ISM02.4	PPB	5/3/2017 10:09 SMPL
I2910-14	MF4E45	225	0.0617 -	ISM02.4	PPB	5/3/2017 10:11 SMPL
I2910-15	MF4E46	385	0.0956 -	ISM02.4	PPB	5/3/2017 10:13 SMPL
I2910-16	MF4E47	488	0.1174 -	ISM02.4	PPB	5/3/2017 10:15 SMPL
I2910-17	MF4E48	326	0.0831 -	ISM02.4	PPB	5/3/2017 10:18 SMPL
CCV025	CCV025	23770	5.0428 -	ISM02.4	PPB	5/3/2017 10:20 SMPL
CCB025	CCB025	51	0.0249 -	ISM02.4	PPB	5/3/2017 10:22 SMPL
I2910-18	MF4E50	272	0.0717 -	ISM02.4	PPB	5/3/2017 10:25 SMPL
I2910-19	MF4E51	299	0.0774 -	ISM02.4	PPB	5/3/2017 10:27 SMPL
I2910-20	MF4E52	810	0.1855 -	ISM02.4	PPB	5/3/2017 10:29 SMPL
I2910-21	MF4E53	827	0.1891 -	ISM02.4	PPB	5/3/2017 10:31 SMPL
PB98661BL	PBS004	82	0.0315 -	ISM02.4	PPB	5/3/2017 10:36 SMPL
I2911-02	MF4E60	284	0.0742 -	ISM02.4	PPB	5/3/2017 10:39 SMPL
I2911-03	MF4E61	158	0.0476 -	ISM02.4	PPB	5/3/2017 10:41 SMPL
I2911-04	MF4E62	212	0.059 -	ISM02.4	PPB	5/3/2017 10:43 SMPL
I2911-05	MF4E63	182	0.0526 -	ISM02.4	PPB	5/3/2017 10:45 SMPL
I2911-06	MF4E64	362	0.0907 -	ISM02.4	PPB	5/3/2017 10:48 SMPL
I2911-07	MF4E32	354	0.089 -	ISM02.4	PPB	5/3/2017 10:50 SMPL
I2911-08	MF4E33	209	0.0584 -	ISM02.4	PPB	5/3/2017 10:52 SMPL
I2911-09	MF4E34	341	0.0863 -	ISM02.4	PPB	5/3/2017 10:55 SMPL
I2911-10	MF4E35	444	0.1081 -	ISM02.4	PPB	5/3/2017 10:57 SMPL
I2911-11	MF4E36	618	0.1449 -	ISM02.4	PPB	5/3/2017 10:59 SMPL
I2911-12	MF4E36D	679	0.1578 -	ISM02.4	PPB	5/3/2017 11:04 SMPL
I2911-13	MF4E36S	12258	2.6074 -	ISM02.4	PPB	5/3/2017 11:06 SMPL
I2911-14	MF4E37	142	0.0442 -	ISM02.4	PPB	5/3/2017 11:09 SMPL
I2911-15	MF4E38	169	0.0499 -	ISM02.4	PPB	5/3/2017 11:11 SMPL
I2911-16	MF4E39	591	0.1392 -	ISM02.4	PPB	5/3/2017 11:13 SMPL
CCV026	CCV026	24401	5.1763 -	ISM02.4	PPB	5/3/2017 11:15 SMPL
CCB026	CCB026	40	0.0226 -	ISM02.4	PPB	5/3/2017 11:18 SMPL
I2911-17	MF4E40	369	0.0922 -	ISM02.4	PPB	5/3/2017 11:20 SMPL
I2911-18	MF4E41	497	0.1193 -	ISM02.4	PPB	5/3/2017 11:22 SMPL
I2911-19	MF4E42	368	0.092 -	ISM02.4	PPB	5/3/2017 11:25 SMPL
I2911-20	MF4E54	318	0.0814 -	ISM02.4	PPB	5/3/2017 11:27 SMPL
I2911-21	MF4E55	265	0.0702 -	ISM02.4	PPB	5/3/2017 11:29 SMPL
CCV027	CCV027	24179	5.1293 -	ISM02.4	PPB	5/3/2017 11:31 SMPL
CCB027	CCB027	78	0.0306 -	ISM02.4	PPB	5/3/2017 11:34 SMPL