



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: AMANDEEP

Parameter: Color

Supervisor Review By: apatel

Run Number: LB87210

Reagent/Standard	Lot/Log #
30 color units	WP54632
25 color units	WP54631
20 color units	WP54630
15 color units	WP54629
5 color units	WP54627
70 color units	WP54637
60 color units	WP54636
50 color units	WP54635
40 color units	WP54634
0 color units	WP54626
10 color units	WP54628

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB87210BL	LB87210BL	1	5.01	5.01	0	05/03/2017	10:00
2	I2942-01	MW-4	1	6.16	6.16	5	05/03/2017	10:05
3	I2942-01DUP	MW-4DUP	1	6.16	6.16	5	05/03/2017	10:10
4	I2942-02	MW-5	1	6.32	6.32	0	05/03/2017	10:17
5	I2943-01	MW-1	1	5.39	5.39	5	05/03/2017	10:25
6	I2943-02	MW-2	1	9.16	9.16	5	05/03/2017	10:34
7	I2943-03	SW-1	10	5.33	5.33	50	05/03/2017	10:40
8	I2944-01	SW-2	10	6.28	6.28	40	05/03/2017	10:46
9	I2944-02	MW-3	1	7.41	7.41	5	05/03/2017	10:52

03-87216

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Color050217

WorkList ID : 98022

Date : 5/2/2017 11:12:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12942-01	Color	Cool 4 deg C	TECT01	D42	MW-4	05/01/2017	SM2120 B
	Water	12942-02	Color	Cool 4 deg C	TECT01	D42	MW-5	05/01/2017	SM2120 B
	Water	12943-01	Color	Cool 4 deg C	TECT01	D32	MW-1	05/01/2017	SM2120 B
	Water	12943-02	Color	Cool 4 deg C	TECT01	D32	MW-2	05/01/2017	SM2120 B
	Water	12943-03	Color	Cool 4 deg C	TECT01	D32	SW-1	05/01/2017	SM2120 B
	Water	12944-01	Color	Cool 4 deg C	TECT01	D51	SW-2	05/01/2017	SM2120 B
	Water	12944-02	Color	Cool 4 deg C	TECT01	D51	MW-3	05/01/2017	SM2120 B

Date/Time 5-3-17 8:45
 Received by: AK
 Relinquished by: AK

Date/Time 5-3-17 3:35
 Received by: IB
 Relinquished by: AK