



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87211

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/02/2017	18:10
2	CAL2	1	Water			7.00	05/02/2017	18:11
3	CAL3	1	Water			10.03	05/02/2017	18:12
4	ICV	1	Water			7.00	05/02/2017	18:13
5	CCV1	1	Water			2.02	05/02/2017	18:15
6	I2949-07	1	Water			6.07	05/02/2017	18:16
7	I2949-07DUP	1	Water			6.08	05/02/2017	18:17
8	CCV2	1	Water			12.02	05/02/2017	18:18

LB87211

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12949

WorkList ID : 98064

Date : 5/2/2017 17:50:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2949-07	pH	Cool 4 deg C	LIR001	D62	TWP-07	05/02/2017	SM 4500-PH B

Date/Time 05/02/17 17:50 PM
 Received by: SR
 Relinquished by: SR

Date/Time 05/02/17 7:00 PM
 Received by: SR
 Relinquished by: SR