



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87213

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/03/2017	11:20
2	CAL2	1	Water			7.01	05/03/2017	11:21
3	CAL3	1	Water			10.03	05/03/2017	11:22
4	ICV	1	Water			7.00	05/03/2017	11:23
5	CCV1	1	Water			2.01	05/03/2017	11:25
6	I2947-21	1	Solid	20.02	20	7.80	05/03/2017	11:26
7	I2947-21DUP	1	Solid	20.04	20	7.81	05/03/2017	11:27
8	I2947-22	1	Solid	20.01	20	6.73	05/03/2017	11:28
9	I2947-23	1	Solid	20.05	20	6.16	05/03/2017	11:30
10	I2947-24	1	Solid	20.02	20	7.38	05/03/2017	11:31
11	I2947-25	1	Solid	20.04	20	8.25	05/03/2017	11:32
12	I2947-26	1	Solid	20.00	20	8.70	05/03/2017	11:33
13	I2949-03	1	Solid	20.02	20	8.43	05/03/2017	11:34
14	I2953-04	1	Solid	20.03	20	8.43	05/03/2017	11:35
15	I2956-03	1	Solid	20.01	20	5.16	05/03/2017	11:36
16	CCV2	1	Water			12.02	05/03/2017	11:37
17	I2956-06	1	Solid	20.04	20	7.11	05/03/2017	11:38
18	I2957-03	1	Solid	20.02	20	7.27	05/03/2017	11:40
19	I2957-06	1	Solid	20.05	20	9.18	05/03/2017	11:41
20	CCV3	1	Water			12.01	05/03/2017	11:42

2387213

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2953 WorkList ID : 98066 Date : 5/3/2017 11:05:38 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2947-21	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-01-COMP	05/01/2017	9045C
	Solid	I2947-22	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-02-COMP	05/01/2017	9045C
	Solid	I2947-23	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-03-COMP	05/02/2017	9045C
	Solid	I2947-24	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-04-COMP	05/01/2017	9045C
	Solid	I2947-25	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-05-COMP	05/01/2017	9045C
	Solid	I2947-26	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-06-COMP	05/02/2017	9045C
	Solid	I2949-03	Corrosivity	Cool 4 deg C	LIRO01	D62	SB-07-COMP	05/02/2017	9045C
	Solid	I2953-04	Corrosivity	Cool 4 deg C	DEWB01	D62	TP7-8-COMP	05/02/2017	9045C
05/05/2017	Solid	I2956-03	Corrosivity	Cool 4 deg C	PSEG01	O22	IP-TP(0-5)	05/02/2017	9045C
05/05/2017	Solid	I2956-06	Corrosivity	Cool 4 deg C	PSEG01	O22	CHERRY-ST-TP(0-5)	05/02/2017	9045C
05/05/2017	Solid	I2957-03	Corrosivity	Cool 4 deg C	PSEG01	O22	RT-3025	05/02/2017	9045C
05/05/2017	Solid	I2957-06	Corrosivity	Cool 4 deg C	PSEG01	O22	RT-2342	05/02/2017	9045C

Date/Time
Received by:
Relinquished by:

05/03/17
J
XR

05/03/17
G1-000
CD
SD

Date/Time
Received by:
Relinquished by:

05/03/17
J
XR

05/03/17
G1-000
CD
SD