


ISM02.4

LB 87 228
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

44692



A= 0.0000e+000

B= 2.2408e-004

C= 3.9440e-002

Rho= 0.9999349

Accept=Accepted

slope
+ intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	-7	0.000	-7					-
0.2	0.200	0.215	0.015	785	0.0 %	785					7
2.5	2.500	2.487	-0.013	10922	0.0 %	10922					-1
5.0	5.000	4.926	-0.074	21808	0.0 %	21808					-1
7.5	7.500	7.479	-0.021	33202	0.0 %	33202					0
10.0	10.000	10.054	0.054	44692	0.0 %	44692					1

ms
5/31/7

LB87228

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		-7	-		0 ISM02.4	PPB	5/3/2017 16:35	Std
0.2 S0.2		785	-		0.2 ISM02.4	PPB	5/3/2017 16:37	Std
2.5 S2.5		10922	-		2.5 ISM02.4	PPB	5/3/2017 16:40	Std
5 S5		21808	-		5 ISM02.4	PPB	5/3/2017 16:42	Std
7.5 S7.5		33202	-		7.5 ISM02.4	PPB	5/3/2017 16:44	Std
10 S10		44692	-		10 ISM02.4	PPB	5/3/2017 16:46	Std
ICV	ICV	16438	3.7229 -		ISM02.4	PPB	5/3/2017 16:49	SMPL
ICB	ICB	-108	0.0152 -		ISM02.4	PPB	5/3/2017 16:51	SMPL
CCV028	CCV028	20988	4.7425 -		ISM02.4	PPB	5/3/2017 16:54	SMPL
CCB028	CCB028	-114	0.0139 -		ISM02.4	PPB	5/3/2017 16:56	SMPL
PB98654BL	PBW003	-42	0.03 -		ISM02.4	PPB	5/3/2017 16:58	SMPL
I2910-01	MF4A11	-49	0.0285 -		ISM02.4	PPB	5/3/2017 17:00	SMPL
I2911-01	MF4A12	-43	0.0298 -		ISM02.4	PPB	5/3/2017 17:03	SMPL
CCV029	CCV029	20879	4.7181 -		ISM02.4	PPB	5/3/2017 17:05	SMPL
CCB029	CCB029	-109	0.015 -		ISM02.4	PPB	5/3/2017 17:07	SMPL