

7470A


LB 87234

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35958



A= 0.0000e+000

B= 2.7692e-004

C= -1.0025e-002

Rho= 0.9999364

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	57	0.000	57				
0.2	0.200	0.202	0.002	767	0.0 %	767				
2.5	2.500	2.452	-0.048	8891	0.0 %	8891				
5.0	5.000	5.023	0.023	18176	0.0 %	18176				
7.5	7.500	7.569	0.069	27369	0.0 %	27369				
10.0	10.000	9.947	-0.053	35958	0.0 %	35958				

LB87234 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 May 2017 08:19:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.00007470A	PPB		04 May 2017 08:24:36	S	Std
0.2 - 1	50.2	767	-	0.20007470A	PPB		04 May 2017 08:26:45	S	Std
2.5 - 1	52.5	8891	-	2.50007470A	PPB		04 May 2017 08:28:53	S	Std
5.0 - 1	55	18176	-	5.00007470A	PPB		04 May 2017 08:31:02	S	Std
7.5 - 1	57.5	27369	-	7.50007470A	PPB		04 May 2017 08:33:09	S	Std
10.0 - 1	510	35958	-	10.00007470A	PPB		04 May 2017 08:35:17	S	Std
ICV41 - 1	ICV41	14053	3.8815	-7470A	PPB		04 May 2017 08:37:55	U	SMPL
ICB41 - 1	ICB41	-57	-0.0258	-7470A	PPB		04 May 2017 08:40:01	U	SMPL
CCV30 - 1	CCV30	17954	4.9618	-7470A	PPB		04 May 2017 08:42:07	U	SMPL
CCB30 - 1	CCB30	-56	-0.0255	-7470A	PPB		04 May 2017 08:44:14	U	SMPL
CRA - 1	CRA	739	0.1946	-7470A	PPB		04 May 2017 08:46:20	U	SMPL
HighStd - 1	HighStd	35650	9.8622	-7470A	PPB		04 May 2017 08:48:26	U	SMPL
ChkStd - 1	ChkStd	24808	6.8598	-7470A	PPB		04 May 2017 08:50:33	U	SMPL
PB98682BL - 1	PBW	-68	-0.0289	-7470A	PPB		04 May 2017 08:52:41	U	SMPL
PB98682BS - 1	LCSW	13137	3.6279	-7470A	PPB		04 May 2017 08:54:48	U	SMPL
PB98682TB - 1	PB98623TB	-28	-0.0178	-7470A	PPB		04 May 2017 08:56:56	U	SMPL
I2932-01 - 1	TP-1	34	-0.0006	-7470A	PPB		04 May 2017 08:59:04	U	SMPL
I2932-02 - 1	TP-2	45	0.0024	-7470A	PPB		04 May 2017 09:01:12	U	SMPL
I2932-03 - 1	TP-3	60	0.0066	-7470A	PPB		04 May 2017 09:03:21	U	SMPL
I2932-04 - 1	TP-4	3	-0.0092	-7470A	PPB		04 May 2017 09:05:27	U	SMPL
CCV31 - 1	CCV31	17792	4.9169	-7470A	PPB		04 May 2017 09:07:34	U	SMPL
CCB31 - 1	CCB31	-62	-0.0272	-7470A	PPB		04 May 2017 09:09:41	U	SMPL
I2932-05 - 1	TP-5	18	-0.0050	-7470A	PPB		04 May 2017 09:11:47	U	SMPL
I2932-06 - 1	TP-6	54	0.0049	-7470A	PPB		04 May 2017 09:13:54	U	SMPL
I2936-06 - 1	TP2-COMP(1-10FT)	48	0.0033	-7470A	PPB		04 May 2017 09:16:01	U	SMPL
I2936-12 - 1	TP3-COMP(0.5-4.5FT)	32	-0.0012	-7470A	PPB		04 May 2017 09:18:08	U	SMPL
I2936-18 - 1	TP6-COMP(0.5-5FT)	106	0.0193	-7470A	PPB		04 May 2017 09:20:16	U	SMPL
I2940-03 - 1	5040-COMP	19	-0.0048	-7470A	PPB		04 May 2017 09:22:24	U	SMPL
I2940-03DUP - 1	5040-COMPDUP	39	0.0008	-7470A	PPB		04 May 2017 09:24:32	U	SMPL
I2940-03MS - 1	5040-COMPMS	13084	3.6132	-7470A	PPB		04 May 2017 09:26:40	U	SMPL
I2940-03MSD - 1	5040-COMPMSD	13048	3.6032	-7470A	PPB		04 May 2017 09:28:49	U	SMPL
CCV32 - 1	CCV32	17851	4.9333	-7470A	PPB		04 May 2017 09:30:56	U	SMPL
CCB32 - 1	CCB32	-78	-0.0316	-7470A	PPB		04 May 2017 09:33:03	U	SMPL
I2940-03LX5 - 1		14	-0.0061	-7470A	PPB		04 May 2017 09:35:10	U	SMPL
I2940-03A - 1		13986	3.8630	-7470A	PPB		04 May 2017 09:37:17	U	SMPL
CCV33 - 1	CCV33	18287	5.0540	-7470A	PPB		04 May 2017 09:39:24	U	SMPL
CCB33 - 1	CCB33	-53	-0.0247	-7470A	PPB		04 May 2017 09:41:30	U	SMPL