



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87236

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/04/2017	09:50
2	CAL2	1	Water			7.00	05/04/2017	09:51
3	CAL3	1	Water			10.03	05/04/2017	09:52
4	ICV	1	Water			7.00	05/04/2017	09:53
5	CCV1	1	Water			2.01	05/04/2017	10:00
6	I2872-02	1	Solid	20.02	20	4.96	05/04/2017	10:01
7	I2969-02	1	Solid	20.03	20	8.39	05/04/2017	10:02
8	I2969-12	1	Solid	20.04	20	8.78	05/04/2017	10:03
9	I2969-12DUP	1	Solid	20.05	20	8.79	05/04/2017	10:05
10	CCV2	1	Water			12.02	05/04/2017	10:06

LB 8723C

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2969 WorkList ID : 98105 Date : 5/4/2017 9:32:30 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2969-02	Corrosivity	Cool 4 deg C	DEWB01	E31	TP4-5COMP	05/01/2017	9045C
	Solid	I2969-12	Corrosivity	Cool 4 deg C	DEWB01	E31	TP1-COMP	05/03/2017	9045C

Date/Time 05/04/17 9:32 AM
Received by: JP
Relinquished by: [Signature]

Date/Time 05/04/17 11:10 AM
Received by: [Signature]
Relinquished by: JP