



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87238

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2183         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2154         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2140         |
| pH 2.00 Buffer                 | W2079         |
| Buffer Solution, PH12 (500ml)  | W2099         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID    | DF | Matrix | Weight (gm) | Volume (ml) | Result (pH) | Anal Date  | Anal Time |
|-----|----------|----|--------|-------------|-------------|-------------|------------|-----------|
| 1   | CAL1     | 1  | Water  |             |             | 4.01        | 05/04/2017 | 09:50     |
| 2   | CAL2     | 1  | Water  |             |             | 7.00        | 05/04/2017 | 09:51     |
| 3   | CAL3     | 1  | Water  |             |             | 10.03       | 05/04/2017 | 09:52     |
| 4   | ICV      | 1  | Water  |             |             | 7.00        | 05/04/2017 | 09:53     |
| 5   | CCV1     | 1  | Water  |             |             | 2.01        | 05/04/2017 | 10:00     |
| 6   | I2872-04 | 1  | Solid  | 20.02       | 20          | 4.96        | 05/04/2017 | 10:01     |
| 7   | CCV2     | 1  | Water  |             |             | 12.02       | 05/04/2017 | 10:06     |