


7471A

LB 87239
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

45960



A= 0.0000e+000

B= 2.1806e-004

C= -2.1084e-002

Rho= 0.9999590

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	241	0.000	241				
0.2	0.200	0.185	-0.015	946	0.0 %	946				
2.5	2.500	2.450	-0.050	11334	0.0 %	11334				
5.0	5.000	5.052	0.052	23263	0.0 %	23263				
7.5	7.500	7.481	-0.019	34403	0.0 %	34403				
10.0	10.000	10.001	0.001	45960	0.0 %	45960				

LB87239 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 04 May 2017 10:10:28				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SC	241	-	0.00007471A	PPB	04 May 2017 10:16:06	S	Std
0.2 - 1	SP-2	946	-	0.20007471A	PPB	04 May 2017 10:18:23	S	Std
2.5 - 1	S2-5	11334	-	2.50007471A	PPB	04 May 2017 10:20:41	S	Std
5.0 - 1	S5	23263	-	5.00007471A	PPB	04 May 2017 10:22:58	S	Std
7.5 - 1	S7-5	34403	-	7.50007471A	PPB	04 May 2017 10:25:15	S	Std
10.0 - 1	S10	45960	-	10.00007471A	PPB	04 May 2017 10:27:32	S	Std
ICV42 - 1	ICV42	17699	3.8383	-7471A	PPB	04 May 2017 10:30:13	U	SMPL
ICB42 - 1	ICB42	153	0.0123	-7471A	PPB	04 May 2017 10:32:28	U	SMPL
CCV34 - 1	CCV34	23165	5.0302	-7471A	PPB	04 May 2017 10:34:44	U	SMPL
CCB34 - 1	CCB34	142	0.0099	-7471A	PPB	04 May 2017 10:36:59	U	SMPL
CRA - 1	CRA	941	0.1841	-7471A	PPB	04 May 2017 10:39:15	U	SMPL
HighStd - 1	HighStd	45760	9.9571	-7471A	PPB	04 May 2017 10:41:30	U	SMPL
ChkStd - 1	ChkStd	31548	6.8581	-7471A	PPB	04 May 2017 10:43:46	U	SMPL
PB98691BL - 1	PBS	3	-0.0204	-7471A	PPB	04 May 2017 10:46:04	U	SMPL
PB98691BS - 1	LCSS	16073	3.4837	-7471A	PPB	04 May 2017 10:48:20	U	SMPL
I2936-05 - 1	TP2-COMP(1-10FT)	7079	1.5225	-7471A	PPB	04 May 2017 10:55:28	U	SMPL
I2936-05DUP - 1	TP2-COMP(1-10FT)DUP	7048	1.5158	-7471A	PPB	04 May 2017 10:57:45	U	SMPL
I2936-05MS - 1	TP2-COMP(1-10FT)MS	24920	5.4129	-7471A	PPB	04 May 2017 11:00:02	U	SMPL
I2936-05MSD - 1	TP2-COMP(1-10FT)MSD	25413	5.5204	-7471A	PPB	04 May 2017 11:02:19	U	SMPL
I2936-11 - 1	TP3-COMP(0.5-4.5FT)	4107	0.8745	-7471A	PPB	04 May 2017 11:04:36	U	SMPL
CCV35 - 1	CCV35	23223	5.0428	-7471A	PPB	04 May 2017 11:06:52	U	SMPL
CCB35 - 1	CCB35	176	0.0173	-7471A	PPB	04 May 2017 11:09:07	U	SMPL
I2936-17 - 1	TP6-COMP(0.5-5FT)	32168	6.9933	-7471A	PPB	04 May 2017 11:11:24	U	SMPL
I2940-01 - 1	5040-COMP	2108	0.4386	-7471A	PPB	04 May 2017 11:13:40	U	SMPL
I2949-01 - 1	SB-07-7.5-8.0	880	0.1708	-7471A	PPB	04 May 2017 11:15:56	U	SMPL
I2953-03 - 1	TP7-8-COMP	56661	12.3342	-7471A	PPB	04 May 2017 11:18:12	U	SMPL
I2956-01 - 1	IP-TP(0-5)	754	0.1433	-7471A	PPB	04 May 2017 11:20:29	U	SMPL
I2956-04 - 1	CHERRY-ST-TP(0-5)	4002	0.8516	-7471A	PPB	04 May 2017 11:22:46	U	SMPL
I2957-01 - 1	RT-3025	965	0.1893	-7471A	PPB	04 May 2017 11:25:03	U	SMPL
I2957-04 - 1	RT-2342	2449	0.5129	-7471A	PPB	04 May 2017 11:27:20	U	SMPL
PB98692BL - 1	PBS	104	0.0016	-7471A	PPB	04 May 2017 11:29:38	U	SMPL
PB98692BS - 1	LCSS	16422	3.5598	-7471A	PPB	04 May 2017 11:31:54	U	SMPL
CCV36 - 1	CCV36	23300	5.0596	-7471A	PPB	04 May 2017 11:34:11	U	SMPL
CCB36 - 1	CCB36	192	0.0208	-7471A	PPB	04 May 2017 11:36:27	U	SMPL
I2947-01 - 1	SB-01-0-2.0	2545	0.5339	-7471A	PPB	04 May 2017 11:38:43	U	SMPL
I2947-01DUP - 1	SB-01-0-2.0DUP	3052	0.6444	-7471A	PPB	04 May 2017 11:40:59	U	SMPL
I2947-01MS - 1	SB-01-0-2.0MS	19385	4.2059	-7471A	PPB	04 May 2017 11:43:16	U	SMPL
I2947-01MSD - 1	SB-01-0-2.0MSD	19653	4.2644	-7471A	PPB	04 May 2017 11:45:32	U	SMPL
I2947-02 - 1	SB-01-7.0-7.5	1271	0.2561	-7471A	PPB	04 May 2017 11:47:49	U	SMPL
I2947-04 - 1	SB-02-0-2.0	3351	0.7096	-7471A	PPB	04 May 2017 11:50:06	U	SMPL
I2947-05 - 1	SB-02-6.0-6.5	462	0.0797	-7471A	PPB	04 May 2017 11:52:23	U	SMPL
I2947-07 - 1	SB-03-0-2.0	570	0.1032	-7471A	PPB	04 May 2017 11:54:41	U	SMPL
I2947-08 - 1	SB-03-8.5-9.0	440	0.0749	-7471A	PPB	04 May 2017 11:56:59	U	SMPL
I2947-10 - 1	SB-04-0-2.0	5594	1.1987	-7471A	PPB	04 May 2017 11:59:15	U	SMPL
CCV37 - 1	CCV37	23388	5.0788	-7471A	PPB	04 May 2017 12:01:31	U	SMPL
CCB37 - 1	CCB37	213	0.0254	-7471A	PPB	04 May 2017 12:03:48	U	SMPL
I2947-11 - 1	SB-04-6.5-7.0	3877	0.8243	-7471A	PPB	04 May 2017 12:06:04	U	SMPL
I2947-12 - 1	SB-04-6.5-7.0-DUP	38239	8.3171	-7471A	PPB	04 May 2017 12:08:21	U	SMPL
I2947-14 - 1	SB-05-0-2.0	14903	3.2286	-7471A	PPB	04 May 2017 12:10:37	U	SMPL
I2947-15 - 1	SB-05-4.0-4.5	2460	0.5153	-7471A	PPB	04 May 2017 12:12:53	U	SMPL
I2947-17 - 1	SB-06-0-2.0	1577	0.3228	-7471A	PPB	04 May 2017 12:15:10	U	SMPL
I2947-18 - 1	SB-06-4.5-5.0	325	0.0498	-7471A	PPB	04 May 2017 12:17:27	U	SMPL
I2947-20 - 1	SB-07-0-2.0	9318	2.0108	-7471A	PPB	04 May 2017 12:19:45	U	SMPL
I2950-01 - 1	SB-8(0-2)20170502	2380	0.4979	-7471A	PPB	04 May 2017 12:22:02	U	SMPL
I2950-02 - 1	SB-8(13-15)20170502	401	0.0664	-7471A	PPB	04 May 2017 12:24:20	U	SMPL
I2950-03 - 1	SB-10(13-15)20170502	500	0.0879	-7471A	PPB	04 May 2017 12:26:36	U	SMPL
CCV38 - 1	CCV38	23432	5.0884	-7471A	PPB	04 May 2017 12:31:45	U	SMPL
CCB38 - 1	CCB38	210	0.0247	-7471A	PPB	04 May 2017 12:34:01	U	SMPL
I2950-04 - 1	SB-10(18-20)20170502	496	0.0871	-7471A	PPB	04 May 2017 12:36:17	U	SMPL
I2950-05 - 1	SB-11(0-2)20170502	1356	0.2746	-7471A	PPB	04 May 2017 12:38:33	U	SMPL
I2950-06 - 1	SB-11(13-15)20170502	582	0.1058	-7471A	PPB	04 May 2017 12:40:50	U	SMPL
I2936-05LX5 - 1		1544	0.3156	-7471A	PPB	04 May 2017 12:43:06	U	SMPL
I2936-05A - 1		24689	5.3625	-7471A	PPB	04 May 2017 12:45:23	U	SMPL
I2947-01LX5 - 1		626	0.1154	-7471A	PPB	04 May 2017 12:47:40	U	SMPL
I2947-01A - 1		20374	4.4216	-7471A	PPB	04 May 2017 12:49:57	U	SMPL
I2953-03DLX5 - 1		11256	2.4333	-7471A	PPB	04 May 2017 12:52:15	U	SMPL
CCV39 - 1	CCV39	23342	5.0688	-7471A	PPB	04 May 2017 12:54:33	U	SMPL
CCB39 - 1	CCB39	217	0.0262	-7471A	PPB	04 May 2017 12:56:50	U	SMPL