



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87248

Temp.: 19.6 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/04/2017	12:30
2	CAL2	1	Water			7.00	05/04/2017	12:31
3	CAL3	1	Water			10.02	05/04/2017	12:32
4	ICV	1	Water			7.00	05/04/2017	12:33
5	CCV1	1	Water			2.02	05/04/2017	12:34
6	I2985-02	1	Water			4.64	05/04/2017	12:35
7	I2985-02DUP	1	Water			4.65	05/04/2017	12:36
8	I2986-01	1	Water			7.08	05/04/2017	12:37
9	CCV2	1	Water			12.02	05/04/2017	12:40

2087248

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12985

WorkList ID : 98114

Date : 5/4/2017 12:22:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2985-02 ✓	pH	Cool 4 deg C	STER16	E31	MAINCELL	05/01/2017	SM 4500-PH B
	Water	I2986-01-17	pH	Cool 4 deg C	STER14	E32	POOLWATER	05/02/2017	SM 4500-PH B

Date/Time 05/04/17 12:22PM
 Received by: Jo
 Relinquished by:

Date/Time 05/04/17 1:35PM
 Received by: CD
 Relinquished by: CD