

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87249

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54662
calibration std. phosphate 0.5 ppm	WP54661
calibration std. phosphate 0.3 ppm	WP54660
calibration std. 0 ppm	WP54657
phosphate ICV std.	WP54664
calibration std. phosphate 0.1 ppm	WP54659
calibration std. phosphate 0.05 ppm	WP54658
Combined reagent	WP54666
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate CCV std.	WP54663
Ascorbic Acid	WP54665

Intercept: -0.0017Slope: 0.6534Regression: 0.999852

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/04/2017	11:21
2	CAL2	0.05	1	50	50	0.032	05/04/2017	11:21
3	CAL3	0.10	1	50	50	0.065	05/04/2017	11:22
4	CAL4	0.30	1	50	50	0.186	05/04/2017	11:22
5	CAL5	0.50	1	50	50	0.329	05/04/2017	11:23
6	CAL6	1.00	1	50	50	0.652	05/04/2017	11:23



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87249

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.319	0.491	05/04/2017	11:24
2	ICB		1	50	50	0.000	0.003	05/04/2017	11:24
3	CCV1	0.50	1	50	50	0.328	0.505	05/04/2017	11:25
4	CCB1		1	50	50	0.000	0.003	05/04/2017	11:25
5	PB98701BL		1	50	50	0.000	0.003	05/04/2017	11:26
6	PB98701BS	0.50	1	50	50	0.314	0.483	05/04/2017	11:26
7	I2901-01		1	50	50	2.953	4.522	05/04/2017	11:27
8	I2901-01DUP		1	50	50	2.879	4.409	05/04/2017	11:27
9	I2901-01MS	0.50	1	50	50	3.071	4.703	05/04/2017	11:28
10	I2901-01MSD	0.50	1	50	50	3.078	4.713	05/04/2017	11:28
11	I2901-01		50	50	50	0.533	0.818	05/04/2017	11:48
12	I2901-01DUP		50	50	50	0.531	0.815	05/04/2017	11:48
13	I2901-01MS	0.50	50	50	50	0.539	0.828	05/04/2017	11:49
14	I2901-01MSD	0.50	50	50	50	0.540	0.829	05/04/2017	11:49
15	CCV2	0.50	1	50	50	0.327	0.503	05/04/2017	11:50
16	CCB2		1	50	50	0.000	0.003	05/04/2017	11:50