



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87253

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/04/2017	14:25
2	CAL2	1	Water			7.00	05/04/2017	14:26
3	CAL3	1	Water			10.02	05/04/2017	14:27
4	ICV	1	Water			7.01	05/04/2017	14:30
5	CCV1	1	Water			2.02	05/04/2017	14:31
6	I2995-01	1	Solid	20.02	20	9.03	05/04/2017	14:32
7	I2995-01DUP	1	Solid	20.04	20	9.04	05/04/2017	14:33
8	I2995-02	1	Solid	20.01	20	9.00	05/04/2017	14:34
9	I2995-03	1	Solid	20.02	20	8.85	05/04/2017	14:35
10	I2995-04	1	Solid	20.03	20	8.87	05/04/2017	14:36
11	I2995-05	1	Solid	20.02	20	8.81	05/04/2017	14:37
12	I2995-06	1	Solid	20.01	20	7.77	05/04/2017	14:38
13	CCV2	1	Water			12.02	05/04/2017	14:40

WORKLIST(Hardcopy Internal Chain)

2387253

WorkList Name : CORROSIONITY I2995

WorkList ID : 98120

Date : 5/4/2017 2:15:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2995-01	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-AOC17-001	05/03/2017	9045C
	Solid	I2995-02	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-AOC17-002	05/03/2017	9045C
	Solid	I2995-03	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-AOC17-003	05/03/2017	9045C
	Solid	I2995-04	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-AOC17-004	05/03/2017	9045C
	Solid	I2995-05	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-AOC17-005	05/03/2017	9045C
	Solid	I2995-06	Corrosivity	Cool 4 deg C	ENTA04	E11	WC-HS-006	05/03/2017	9045C

Date/Time 05/04/17 2:15 PM
 Received by: SP
 Relinquished by: [Signature]

Date/Time 05/04/17 2:00 PM
 Received by: [Signature]
 Relinquished by: 20