

7471A


LB87255

INSTRUMENT ID: CV4

Linear

 μ Abs.:

44636



A= 0.0000e+000

B= 2.2406e-004

C= -2.8038e-003

Rho= 0.999979

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	48	0.000	48				
0.2	0.200	0.201	0.001	911	0.0 %	911				
2.5	2.500	2.489	-0.011	11121	0.0 %	11121				
5.0	5.000	4.993	-0.007	22296	0.0 %	22296				
7.5	7.500	7.510	0.010	33531	0.0 %	33531				
10.0	10.000	9.999	-0.001	44636	0.0 %	44636				

LB87255 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 04 May 2017 14:59:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	7471A	PPB	04 May 2017 15:12:16	S	Std
0.2 - 1	50.2	911	-	0.2000	7471A	PPB	04 May 2017 15:14:34	S	Std
2.5 - 1	52.5	11121	-	2.5000	7471A	PPB	04 May 2017 15:16:52	S	Std
5.0 - 1	55	22296	-	5.0000	7471A	PPB	04 May 2017 15:19:10	S	Std
7.5 - 1	57.5	33531	-	7.5000	7471A	PPB	04 May 2017 15:21:27	S	Std
10.0 - 1	510	44636	-	10.0000	7471A	PPB	04 May 2017 15:23:43	S	Std
ICV43 - 1	ICV43	17329	3.8800	-	7471A	PPB	04 May 2017 15:26:22	U	SMPL
ICB43 - 1	ICB43	-34	-0.0104	-	7471A	PPB	04 May 2017 15:28:37	U	SMPL
CCV40 - 1	CCV40	21874	4.8984	-	7471A	PPB	04 May 2017 15:30:53	U	SMPL
CCB40 - 1	CCB40	-38	-0.0113	-	7471A	PPB	04 May 2017 15:33:09	U	SMPL
CRA - 1	CRA	862	0.1903	-	7471A	PPB	04 May 2017 15:35:25	U	SMPL
HighStd - 1	HighStd	44261	9.9145	-	7471A	PPB	04 May 2017 15:37:40	U	SMPL
ChkStd - 1	ChkStd	30170	6.7572	-	7471A	PPB	04 May 2017 15:39:57	U	SMPL
PB98709BL - 1	PBS	-55	-0.0151	-	7471A	PPB	04 May 2017 15:42:13	U	SMPL
PB98709BS - 1	LCSS	14657	3.2813	-	7471A	PPB	04 May 2017 15:44:30	U	SMPL
I2859-03 - 1	UT-3	1805	0.4016	-	7471A	PPB	04 May 2017 15:46:47	U	SMPL
I2969-01 - 1	TP4-5COMP	1922	0.4278	-	7471A	PPB	04 May 2017 15:49:04	U	SMPL
I2969-11 - 1	TP1-COMP	7490	1.6754	-	7471A	PPB	04 May 2017 15:51:21	U	SMPL
I2970-01 - 1	E2-U10-BASE	55683	12.4737	-	7471A	PPB	04 May 2017 15:53:39	U	SMPL
I2970-02 - 1	E2-U10-ESW	4136	0.9239	-	7471A	PPB	04 May 2017 15:55:55	U	SMPL
CCV41 - 1	CCV41	21963	4.9183	-	7471A	PPB	04 May 2017 15:58:11	U	SMPL
CCB41 - 1	CCB41	-22	-0.0077	-	7471A	PPB	04 May 2017 16:00:27	U	SMPL
I2970-03 - 1	E2-UV10-ESW	9361	2.0947	-	7471A	PPB	04 May 2017 16:02:43	U	SMPL
I2970-06 - 1	E2-U7-BASE	16401	3.6721	-	7471A	PPB	04 May 2017 16:04:59	U	SMPL
I2977-01 - 1	IM-TS-EME-12	523	0.1144	-	7471A	PPB	04 May 2017 16:07:15	U	SMPL
I2977-01DUP - 1	IM-TS-EME-12DUP	561	0.1229	-	7471A	PPB	04 May 2017 16:09:31	U	SMPL
I2977-01MS - 1	IM-TS-EME-12MS	16095	3.6035	-	7471A	PPB	04 May 2017 16:11:48	U	SMPL
I2977-01MSD - 1	IM-TS-EME-12MSD	16525	3.6999	-	7471A	PPB	04 May 2017 16:14:05	U	SMPL
I2977-01LX5 - 1		142	0.0290	-	7471A	PPB	04 May 2017 16:18:54	U	SMPL
I2977-01A - 1		18734	4.1948	-	7471A	PPB	04 May 2017 16:21:12	U	SMPL
I2970-01DLX2 - 1		28253	6.3277	-	7471A	PPB	04 May 2017 16:23:30	U	SMPL
CCV42 - 1	CCV42	22006	4.9279	-	7471A	PPB	04 May 2017 16:25:46	U	SMPL
CCB42 - 1	CCB42	-33	-0.0102	-	7471A	PPB	04 May 2017 16:28:02	U	SMPL