



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87256

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/04/2017	16:10
2	CAL2	1	Water			7.01	05/04/2017	16:11
3	CAL3	1	Water			10.03	05/04/2017	16:12
4	ICV	1	Water			7.01	05/04/2017	16:15
5	CCV1	1	Water			2.01	05/04/2017	16:17
6	I3004-01	1	Solid	20.02	20	5.18	05/04/2017	16:18
7	I3004-01DUP	1	Solid	20.04	20	5.19	05/04/2017	16:19
8	I3007-01	1	Solid	20.01	20	5.05	05/04/2017	16:20
9	CCV2	1	Water			12.01	05/04/2017	16:21

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3004

WorkList ID : 98121

Date : 5/4/2017 3:44:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3004-01-A	pH	Cool 4 deg C	TULL02	E31	EME-H	05/04/2017	9045C
	Solid	I3007-01-A	pH	Cool 4 deg C	TULL02	E31	EME-BW	05/04/2017	9045C

Date/Time 05/04/17 3:44:43 PM
 Received by: JP
 Relinquished by: JP

Date/Time 05/04/17 3:44:43 PM
 Received by: JP
 Relinquished by: JP