

LB 87283


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33721



A= 0.0000e+000

B= 2.9709e-004

C= 3.2804e-002

Rho= 0.9999398

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	-54	0.000	-54				
0.2	0.200	0.217	0.017	619	0.0 %	619				
2.5	2.500	2.516	0.016	8359	0.0 %	8359				
5.0	5.000	4.923	-0.077	16462	0.0 %	16462				
7.5	7.500	7.476	-0.024	25054	0.0 %	25054				
10.0	10.000	10.051	0.051	33721	0.0 %	33721				

LB87283 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 May 2017 15:32:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-54	-	0.0000	245.1	PPB	05 May 2017 15:50:43	S	Std
0.2 - 1	50.2	619	-	0.2000	245.1	PPB	05 May 2017 15:52:51	S	Std
2.5 - 1	52.5	8359	-	2.5000	245.1	PPB	05 May 2017 15:54:59	S	Std
5.0 - 1	55	16462	-	5.0000	245.1	PPB	05 May 2017 15:57:08	S	Std
7.5 - 1	57.5	25054	-	7.5000	245.1	PPB	05 May 2017 15:59:15	S	Std
10.0 - 1	510	33721	-	10.0000	245.1	PPB	05 May 2017 16:01:22	S	Std
ICV45 - 1	ICV45	13106	3.9264	-245.1		PPB	05 May 2017 16:09:17	U	SMPL
ICB45 - 1	ICB45	3	0.0337	-245.1		PPB	05 May 2017 16:11:24	U	SMPL
CCV51 - 1	CCV51	16158	4.8331	-245.1		PPB	05 May 2017 16:13:30	U	SMPL
CCB51 - 1	CCB51	-15	0.0283	-245.1		PPB	05 May 2017 16:15:37	U	SMPL
CRA - 1	CRA	683	0.2357	-245.1		PPB	05 May 2017 16:17:43	U	SMPL
HighStd - 1	HighStd	33234	9.9062	-245.1		PPB	05 May 2017 16:19:50	U	SMPL
ChkStd - 1	ChkStd	22905	6.8376	-245.1		PPB	05 May 2017 16:21:57	U	SMPL
PB98712BL - 1	PBW	-24	0.0257	-245.1		PPB	05 May 2017 16:24:04	U	SMPL
PB98712BS - 1	LCSW	13315	3.9885	-245.1		PPB	05 May 2017 16:26:12	U	SMPL
I2949-07 - 1	TWP-07	472	0.1730	-245.1		PPB	05 May 2017 16:28:19	U	SMPL
I2968-01 - 1	COMP	595	0.2096	-245.1		PPB	05 May 2017 16:30:28	U	SMPL
I2968-01DUP - 1	COMPDUP	576	0.2039	-245.1		PPB	05 May 2017 16:32:36	U	SMPL
I2968-01MS - 1	COMPMS	11189	3.3569	-245.1		PPB	05 May 2017 16:37:07	U	SMPL
I2968-01MSD - 1	COMPMSD	11206	3.3620	-245.1		PPB	05 May 2017 16:39:13	U	SMPL
CCV52 - 1	CCV52	16104	4.8171	-245.1		PPB	05 May 2017 16:41:21	U	SMPL
CCB52 - 1	CCB52	-115	-0.0014	-245.1		PPB	05 May 2017 16:43:27	U	SMPL
I3019-01 - 1	KITCHEN-FAUCET-PM	-168	-0.0171	-245.1		PPB	05 May 2017 16:45:34	U	SMPL
I3019-02 - 1	KITCHEN-FAUCET-AM	-132	-0.0064	-245.1		PPB	05 May 2017 16:47:41	U	SMPL
I2968-01LX5 - 1		41	0.0450	-245.1		PPB	05 May 2017 16:49:48	U	SMPL
I2968-01A - 1		12073	3.6195	-245.1		PPB	05 May 2017 16:51:55	U	SMPL
CCV53 - 1	CCV53	16027	4.7942	-245.1		PPB	05 May 2017 16:54:02	U	SMPL
CCB53 - 1	CCB53	-47	0.0188	-245.1		PPB	05 May 2017 16:56:10	U	SMPL