



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87288

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/05/2017	16:10
2	CAL2	1	Water			7.00	05/05/2017	16:11
3	CAL3	1	Water			10.03	05/05/2017	16:12
4	ICV	1	Water			6.99	05/05/2017	16:13
5	CCV1	1	Water			2.01	05/05/2017	16:15
6	I3022-02	1	Solid	20.02	20	7.15	05/05/2017	16:16
7	I3027-11	1	Solid	20.04	20	7.26	05/05/2017	16:17
8	I3027-13	1	Solid	20.01	20	6.66	05/05/2017	16:19
9	I3027-15	1	Solid	20.03	20	8.11	05/05/2017	16:20
10	I3031-07	1	Solid	20.05	20	7.94	05/05/2017	16:21
11	I3031-09	1	Solid	20.04	20	7.35	05/05/2017	16:22
12	I3034-08	1	Solid	20.02	20	7.43	05/05/2017	16:23
13	I3034-08DUP	1	Solid	20.03	20	7.44	05/05/2017	16:25
14	CCV2	1	Water			12.01	05/05/2017	16:26

WORKLIST(Hardcopy Internal Chain)

1387288

WorkList Name : CORROSIVITY I3031

WorkList ID : 98175

Date : 5/5/2017 3:58:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3022-02	Corrosivity	Cool 4 deg C	VIAS01	E32	WC-COMP-15	05/04/2017	9045C
	Solid	I3027-11	Corrosivity	Cool 4 deg C	DEWB01	F31	TP-1-2-3-COMP	05/05/2017	9045C
	Solid	I3027-13	Corrosivity	Cool 4 deg C	DEWB01	F31	TP-4-5-6-COMP	05/05/2017	9045C
	Solid	I3027-15	Corrosivity	Cool 4 deg C	DEWB01	F31	TP-7-8-9-COMP	05/05/2017	9045C
	Solid	I3031-07	Corrosivity	Cool 4 deg C	DEWB01	F21	TP-2COMP	05/05/2017	9045C
	Solid	I3031-09	Corrosivity	Cool 4 deg C	DEWB01	F21	TP-1-3COMP	05/05/2017	9045C
05/10/2017	Solid	I3034-08	Corrosivity	Cool 4 deg C	PSEG01	F31	DRUM-1-6-COMP	05/05/2017	9045C

Date/Time 05/05/17 3:58 PM
 Received by: 71
 Relinquished by: CL

Date/Time 05/05/17 7:00 PM
 Received by: JR
 Relinquished by: JP