



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87290

Temp.: 22.4 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/05/2017	17:00
2	CAL2	1	Water			7.00	05/05/2017	17:01
3	CAL3	1	Water			10.02	05/05/2017	17:02
4	ICV	1	Water			6.99	05/05/2017	17:03
5	CCV1	1	Water			2.01	05/05/2017	17:04
6	I3033-05	1	Water			5.39	05/05/2017	17:05
7	I3035-07	1	Water			6.62	05/05/2017	17:06
8	I3035-10	1	Water			6.88	05/05/2017	17:07
9	I3035-10DUP	1	Water			6.89	05/05/2017	17:09
10	CCV2	1	Water			12.01	05/05/2017	17:10

WORKLIST(Hardcopy Internal Chain)

2587290

WorkList Name : CORROSIONITY I3034

WorkList ID : 98178

Date : 5/5/2017 4:01:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/05/2017	Water	I3033-05	Corrosivity	Cool 4 deg C	PSEG01	F31	OILY-WATER-COMP	05/05/2017	9040C
	Water	I3035-07	Corrosivity	Cool 4 deg C	PSEG01	F32	DRUM-COMP-(1-5)	05/05/2017	9040C
	Water	I3035-10	Corrosivity	Cool 4 deg C	PSEG01	F32	TANK-1	05/05/2017	9040C

Date/Time 05/05/17 4:00 PM
 Received by: SR
 Relinquished by: SR

Date/Time 05/05/17 6:30 PM
 Received by: SR
 Relinquished by: SR