


7470A

LB 87298
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34887



A= 0.0000e+000

B= 2.8629e-004

C= -1.2608e-003

Rho= 0.9996594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	108	0.000	108				
0.2	0.200	0.154	-0.046	544	0.0 %	544				
2.5	2.500	2.615	0.115	9137	0.0 %	9137				
5.0	5.000	4.824	-0.176	16853	0.0 %	16853				
7.5	7.500	7.591	0.091	26519	0.0 %	26519				
10.0	10.000	9.987	-0.013	34887	0.0 %	34887				

LB87298 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 May 2017 09:52:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	108	-	0.0000	7470A	PPB	08 May 2017 10:55:34	S	Std
0.2 - 1	50.2	544	-	0.2000	7470A	PPB	08 May 2017 11:00:10	S	Std
2.5 - 1	52.5	9137	-	2.5000	7470A	PPB	08 May 2017 11:02:18	S	Std
5.0 - 1	55	16853	-	5.0000	7470A	PPB	08 May 2017 11:04:26	S	Std
7.5 - 1	57.5	26519	-	7.5000	7470A	PPB	08 May 2017 11:06:34	S	Std
10.0 - 1	58	34887	-	10.0000	7470A	PPB	08 May 2017 11:08:41	S	Std
ICV46 - 1	ICV46	13744	3.9336	-	7470A	PPB	08 May 2017 11:11:27	U	SMPL
ICB46 - 1	ICB46	-145	-0.0428	-	7470A	PPB	08 May 2017 11:13:33	U	SMPL
CCV54 - 1	CCV54	17192	4.9207	-	7470A	PPB	08 May 2017 11:15:40	U	SMPL
CCB54 - 1	CCB54	6	0.0005	-	7470A	PPB	08 May 2017 11:17:46	U	SMPL
CRA - 1	CRA	612	0.1740	-	7470A	PPB	08 May 2017 11:23:08	U	SMPL
HighStd - 1	HighStd	35608	10.1931	-	7470A	PPB	08 May 2017 11:25:14	U	SMPL
ChkStd - 1	ChkStd	24866	7.1177	-	7470A	PPB	08 May 2017 11:27:21	U	SMPL
PB98749BL - 1	PBW	-180	-0.0528	-	7470A	PPB	08 May 2017 11:29:29	U	SMPL
PB98749BS - 1	LCSW	13346	3.8196	-	7470A	PPB	08 May 2017 11:31:37	U	SMPL
PB98749TB - 1	PB98716TB	-122	-0.0362	-	7470A	PPB	08 May 2017 11:33:44	U	SMPL
I2995-01 - 1	WC-AOC17-001	-66	-0.0202	-	7470A	PPB	08 May 2017 11:35:52	U	SMPL
I2995-01DUP - 1	WC-AOC17-001DUP	-61	-0.0187	-	7470A	PPB	08 May 2017 11:38:00	U	SMPL
I2995-01MS - 1	WC-AOC17-001MS	13052	3.7354	-	7470A	PPB	08 May 2017 11:40:09	U	SMPL
I2995-01MSD - 1	WC-AOC17-001MSD	13154	3.7646	-	7470A	PPB	08 May 2017 11:42:16	U	SMPL
CCV55 - 1	CCV55	17704	5.0673	-	7470A	PPB	08 May 2017 11:44:23	U	SMPL
CCB55 - 1	CCB55	-140	-0.0413	-	7470A	PPB	08 May 2017 11:46:30	U	SMPL
I2995-02 - 1	WC-AOC17-002	-43	-0.0136	-	7470A	PPB	08 May 2017 11:48:36	U	SMPL
I2995-03 - 1	WC-AOC17-003	-58	-0.0179	-	7470A	PPB	08 May 2017 11:50:43	U	SMPL
I2995-04 - 1	WC-AOC17-004	-56	-0.0173	-	7470A	PPB	08 May 2017 11:52:50	U	SMPL
I2995-05 - 1	WC-AOC17-005	-45	-0.0141	-	7470A	PPB	08 May 2017 11:54:57	U	SMPL
I2995-06 - 1	WC-HS-006	-40	-0.0127	-	7470A	PPB	08 May 2017 11:57:05	U	SMPL
I2995-01LX5 - 1		51	0.0133	-	7470A	PPB	08 May 2017 12:04:18	U	SMPL
I2995-01A - 1		14005	4.0083	-	7470A	PPB	08 May 2017 12:06:25	U	SMPL
CCV56 - 1	CCV56	17224	4.9299	-	7470A	PPB	08 May 2017 12:08:33	U	SMPL
CCB56 - 1	CCB56	-138	-0.0408	-	7470A	PPB	08 May 2017 12:10:42	U	SMPL