

LB 87302

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

5/8/2017 9:41

Test: Phenolics- *ngk*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.992	0.0	0.109	
ICB1	0.001	0.0	0.002	
CCV1	0.992	0.0	0.109	
CCB1	-0.002	0.0	0.002	
PB98752BL	0.001	0.0	0.002	
PB98752BS	0.990	0.0	0.109	
I2969-01	0.001	0.0	0.002	
I2969-01DUP	0.002	0.0	0.002	
I2969-01MS	0.932	0.0	0.102	
I2969-01MSD	0.928	0.0	0.102	
I2969-11	-0.001	0.0	0.002	
CCV2	0.985	0.0	0.108	
CCB2	0.002	0.0	0.002	

N	13
Mean	0.448
SD	0.5033
CV%	112.41

CB87302

Aquakem v. 7.2AQ1

Results from time period:

Mon May 08 09:03:38 2017

Mon May 08 09:37:44 2017

1A/C

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0498	mg/l	5/8/2017 9:03:38	
0.1PPM	A	Phenolics-	P	0.1045	mg/l	5/8/2017 9:03:39	
0.25PPM	A	Phenolics-	P	0.2546	mg/l	5/8/2017 9:03:40	
0.50PPM	A	Phenolics-	P	0.4964	mg/l	5/8/2017 9:03:41	
1.0PPM	A	Phenolics-	P	0.9894	mg/l	5/8/2017 9:03:42	
2.0PPM	A	Phenolics-	P	2.0054	mg/l	5/8/2017 9:03:43	
ICV1	S	Phenolics-	P	0.9921	mg/l	5/8/2017 9:33:45	
ICB1	S	Phenolics-	P	0.0011	mg/l	5/8/2017 9:33:46	
CCV1	S	Phenolics-	P	0.9918	mg/l	5/8/2017 9:33:47	
CCB1	S	Phenolics-	P	-0.0024	mg/l	5/8/2017 9:33:48	
PB98752BL	S	Phenolics-	P	0.0009	mg/l	5/8/2017 9:33:49	
PB98752BS	S	Phenolics-	P	0.9901	mg/l	5/8/2017 9:33:50	
I2969-01	S	Phenolics-	P	0.0007	mg/l	5/8/2017 9:33:51	
I2969-01DUP	S	Phenolics-	P	0.0016	mg/l	5/8/2017 9:33:52	
I2969-01MS	S	Phenolics-	P	0.9319	mg/l	5/8/2017 9:33:53	
I2969-01MSD	S	Phenolics-	P	0.9276	mg/l	5/8/2017 9:33:54	
I2969-11	S	Phenolics-	P	-0.0009	mg/l	5/8/2017 9:33:55	
CCV2	S	Phenolics-	P	0.9846	mg/l	5/8/2017 9:33:56	
CCB2	S	Phenolics-	P	0.0019	mg/l	5/8/2017 9:37:43	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/8/2017 9:07


Test Phenolics-

Accepted 5/8/2017 9:07

Factor 9.288
Bias 0.002

Coeff. of det. 0.999929

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.007	0.0498	0.0500	
2	Phenol-2	0.013	0.1045	0.1000	
3	Phenol-2	0.029	0.2546	0.2500	
4	Phenol-2	0.055	0.4964	0.5000	
5	Phenol-2	0.108	0.9894	1.0000	
6	Phenol-2	0.218	2.0054	2.0000	