


7471A

LB87305
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45959



A= 0.0000e+000

B= 2.1688e-004

C= 1.8206e-002

Rho= 0.9999749

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	0	0.000	0				
0.2	0.200	0.209	0.009	882	0.0 %	882				
2.5	2.500	2.459	-0.041	11253	0.0 %	11253				
5.0	5.000	4.988	-0.012	22913	0.0 %	22913				
7.5	7.500	7.540	0.040	34683	0.0 %	34683				
10.0	10.000	9.986	-0.014	45959	0.0 %	45959				

LB87305 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 08 May 2017 13:15:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7471A	PPB	08 May 2017 13:59:16	S	Std
0.2 - 1	50.2	882	-	0.2000	7471A	PPB	08 May 2017 14:01:34	S	Std
2.5 - 1	52.5	11253	-	2.5000	7471A	PPB	08 May 2017 14:03:51	S	Std
5.0 - 1	55	22913	-	5.0000	7471A	PPB	08 May 2017 14:06:08	S	Std
7.5 - 1	57.5	34683	-	7.5000	7471A	PPB	08 May 2017 14:11:35	S	Std
10.0 - 1	510	45959	-	10.0000	7471A	PPB	08 May 2017 14:16:32	S	Std
ICV47 - 1	ICV47	17553	3.8251	-7471A	PPB	08 May 2017 14:21:59	U	SMPL	
ICB47 - 1	ICB47	-87	-0.0007	-7471A	PPB	08 May 2017 14:24:14	U	SMPL	
CCV57 - 1	CCV57	22796	4.9622	-7471A	PPB	08 May 2017 14:26:30	U	SMPL	
CCB57 - 1	CCB57	-97	-0.0028	-7471A	PPB	08 May 2017 14:28:46	U	SMPL	
CRA - 1	CRA	858	0.2043	-7471A	PPB	08 May 2017 14:31:02	U	SMPL	
HighStd - 1	HighStd	47068	10.2263	-7471A	PPB	08 May 2017 14:33:18	U	SMPL	
ChkStd - 1	ChkStd	31495	6.8488	-7471A	PPB	08 May 2017 14:35:34	U	SMPL	
PB98750BL - 1	PBS	-134	-0.0109	-7471A	PPB	08 May 2017 14:37:51	U	SMPL	
PB98750BS - 1	LCSS	17672	3.8509	-7471A	PPB	08 May 2017 14:43:00	U	SMPL	
I2976-01 - 1	P001-PC001-01	2141	0.4825	-7471A	PPB	08 May 2017 14:45:16	U	SMPL	
I2976-01DUP - 1	P001-PC001-01DUP	2167	0.4882	-7471A	PPB	08 May 2017 14:47:34	U	SMPL	
I2976-02 - 1	I2976-01MS	20337	4.4289	-7471A	PPB	08 May 2017 14:52:25	U	SMPL	
I2976-03 - 1	I2976-01MSD	19146	4.1706	-7471A	PPB	08 May 2017 14:54:43	U	SMPL	
I2976-04 - 1	P001-PC001-02	1638	0.3735	-7471A	PPB	08 May 2017 14:56:59	U	SMPL	
CCV58 - 1	CCV58	23003	5.0071	-7471A	PPB	08 May 2017 14:59:15	U	SMPL	
CCB58 - 1	CCB58	-77	0.0015	-7471A	PPB	08 May 2017 15:01:31	U	SMPL	
I2976-05 - 1	P001-TANK001-01	34193	7.4340	-7471A	PPB	08 May 2017 15:03:47	U	SMPL	
I2976-09 - 1	P001-TANK002-01	7968	1.7463	-7471A	PPB	08 May 2017 15:06:03	U	SMPL	
I2976-10 - 1	P001-TANK003-01	20367	4.4354	-7471A	PPB	08 May 2017 15:08:20	U	SMPL	
I2976-11 - 1	I2976-10MS	38178	8.2982	-7471A	PPB	08 May 2017 15:10:36	U	SMPL	
I2976-12 - 1	I2976-10MSD	37899	8.2377	-7471A	PPB	08 May 2017 15:12:53	U	SMPL	
I2976-13 - 1	P001-TANK003-02	16271	3.5470	-7471A	PPB	08 May 2017 15:15:10	U	SMPL	
I3027-10 - 1	TP-1-2-3-COMP	1962	0.4437	-7471A	PPB	08 May 2017 15:17:26	U	SMPL	
I3027-12 - 1	TP-4-5-6-COMP	4978	1.0978	-7471A	PPB	08 May 2017 15:19:44	U	SMPL	
I3027-14 - 1	TP-7-8-9-COMP	875	0.2080	-7471A	PPB	08 May 2017 15:22:02	U	SMPL	
I3031-06 - 1	TP-2COMP	1297	0.2995	-7471A	PPB	08 May 2017 15:24:18	U	SMPL	
CCV59 - 1	CCV59	23191	5.0479	-7471A	PPB	08 May 2017 15:26:35	U	SMPL	
CCB59 - 1	CCB59	-52	0.0069	-7471A	PPB	08 May 2017 15:28:51	U	SMPL	
I3031-08 - 1	TP-1-3COMP	1201	0.2787	-7471A	PPB	08 May 2017 15:31:07	U	SMPL	
I3034-01 - 1	DRUM-1-6-COMP	1316	0.3036	-7471A	PPB	08 May 2017 15:33:24	U	SMPL	
PB98751BL - 1	PBS	-26	0.0126	-7471A	PPB	08 May 2017 15:35:40	U	SMPL	
PB98751BS - 1	LCSS	17523	3.8186	-7471A	PPB	08 May 2017 15:37:57	U	SMPL	
2996-01 - 1	SS-01	4926	1.0866	-7471A	PPB	08 May 2017 15:40:14	U	SMPL	
2996-01DUP - 1	SS-01DUP	5768	1.2692	-7471A	PPB	08 May 2017 15:42:30	U	SMPL	
2996-01MS - 1	SS-01MS	23607	5.1381	-7471A	PPB	08 May 2017 15:46:51	U	SMPL	
2996-01MSD - 1	SS-01MSD	23758	5.1708	-7471A	PPB	08 May 2017 15:49:08	U	SMPL	
3004-01 - 1	EME-H	2300	0.5170	-7471A	PPB	08 May 2017 15:51:26	U	SMPL	
3007-01 - 1	EME-BW	3276	0.7287	-7471A	PPB	08 May 2017 15:53:42	U	SMPL	
CCV60 - 1	CCV60	23133	5.0353	-7471A	PPB	08 May 2017 15:55:58	U	SMPL	
CCB60 - 1	CCB60	-56	0.0061	-7471A	PPB	08 May 2017 15:58:15	U	SMPL	
3010-01 - 1	S-1	726	0.1757	-7471A	PPB	08 May 2017 16:00:31	U	SMPL	
3010-03 - 1	S-2	534	0.1340	-7471A	PPB	08 May 2017 16:02:48	U	SMPL	
3011-01 - 1	BP-1	1232	0.2854	-7471A	PPB	08 May 2017 16:05:04	U	SMPL	
3013-01 - 1	E2-V9-BASE	6915	1.5179	-7471A	PPB	08 May 2017 16:07:20	U	SMPL	
3013-02 - 1	E2-U9-BASE	22085	4.8080	-7471A	PPB	08 May 2017 16:09:37	U	SMPL	
3013-03 - 1	E2-UV89-WSW	92877	20.1613	-7471A	PPB	08 May 2017 16:11:54	U	SMPL	
3013-04 - 1	E2-U8-BASE	56106	12.1864	-7471A	PPB	08 May 2017 16:14:11	U	SMPL	
3013-05 - 1	E2-V7-WSW	34083	7.4101	-7471A	PPB	08 May 2017 16:16:30	U	SMPL	
3013-06 - 1	E2-CONCRETE-8-WEST	5337	1.1757	-7471A	PPB	08 May 2017 16:18:48	U	SMPL	
3013-07 - 1	E2-CONCRETE-9-EAST	13361	2.9159	-7471A	PPB	08 May 2017 16:21:04	U	SMPL	
CCV61 - 1	CCV61	23137	5.0361	-7471A	PPB	08 May 2017 16:23:21	U	SMPL	
CCB61 - 1	CCB61	-56	0.0061	-7471A	PPB	08 May 2017 16:25:37	U	SMPL	
3022-01 - 1	WC-COMP-15	4913	1.0837	-7471A	PPB	08 May 2017 16:27:54	U	SMPL	
3024-01 - 1	WETLAND-NYSS	504	0.1275	-7471A	PPB	08 May 2017 16:30:10	U	SMPL	
2976-01LX5 - 1		469	0.1199	-7471A	PPB	08 May 2017 16:32:26	U	SMPL	
2976-01A - 1		18731	4.0806	-7471A	PPB	08 May 2017 16:34:43	U	SMPL	
2996-01LX5 - 1		886	0.2104	-7471A	PPB	08 May 2017 16:36:59	U	SMPL	
2996-01A - 1		22577	4.9147	-7471A	PPB	08 May 2017 16:39:16	U	SMPL	
3013-03DLX5 - 1		18416	4.0123	-7471A	PPB	08 May 2017 16:41:32	U	SMPL	
3013-04DLX2 - 1		28609	6.2229	-7471A	PPB	08 May 2017 16:43:50	U	SMPL	
CCV62 - 1	CCV62	23286	5.0685	-7471A	PPB	08 May 2017 16:46:07	U	SMPL	
CCB62 - 1	CCB62	-49	0.0076	-7471A	PPB	08 May 2017 16:48:24	U	SMPL	