



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87309

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	05/08/2017	16:20
2	I2949-07		1	>100.0	>212.0	05/08/2017	16:50
3	I2949-07DUP		1	>100.0	>212.0	05/08/2017	17:10
4	I3033-05		1	>100.0	>212.0	05/08/2017	17:30
5	I3035-07		1	>100.0	>212.0	05/08/2017	17:50
6	I3035-10		1	>100.0	>212.0	05/08/2017	18:15

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 87309

WorkList Name : FP5817

WorkList ID : 98229

Date : 5/8/2017 3:58:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12949-07 H	Flash Point	Cool 4 deg C	LIR001	D62	TWP-07	05/02/2017	1010A
05/10/2017	Water	13033-05 C	Flash Point	Cool 4 deg C	PSEG01	F31	OILY-WATER-COMP	05/05/2017	1010A
	Water	13035-07 C	Flash Point	Cool 4 deg C	PSEG01	F32	DRUM-COMP-(1-5)	05/05/2017	1010A
	Water	13035-10 C	Flash Point	Cool 4 deg C	PSEG01	F32	TANK-265563	05/05/2017	1010A

Date/Time

5/8/17 4:00 PM

Received by:

KS

Relinquished by:

CP

Date/Time

5/8/17 18:20

Received by:

CP

Relinquished by:

KS