



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87311

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/08/2017	15:00
2	CAL2	1	Water			7.02	05/08/2017	15:02
3	CAL3	1	Water			10.04	05/08/2017	15:03
4	ICV	1	Water			6.99	05/08/2017	15:05
5	CCV1	1	Water			2.01	05/08/2017	15:06
6	I3040-05	1	Solid	20.03	20	8.16	05/08/2017	15:08
7	I3040-05DUP	1	Solid	20.05	20	8.17	05/08/2017	15:09
8	I3048-04	1	Solid	20.01	20	7.68	05/08/2017	15:10
9	I3049-04	1	Solid	20.03	20	8.85	05/08/2017	15:11
10	CCV2	1	Water			12.02	05/08/2017	15:12

2B87311

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3040 WorkList ID : 98224 Date : 5/8/2017 1:55:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3040-05	Corrosivity	Cool 4 deg C	DEWB01	F51	TP-4-5-COMP	05/08/2017	9045C
05/11/2017	Solid	I3048-04	Corrosivity	Cool 4 deg C	PSEG01	F31	RO-4923	05/08/2017	9045C
05/11/2017	Solid	I3049-04	Corrosivity	Cool 4 deg C	PSEG01	F31	DRUM-M4935	05/08/2017	9045C

Date/Time 05/08/17 6:00M
Received by: 
Relinquished by: SP

Date/Time 05/08/17 11:55am
Received by: 
Relinquished by: 