



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87325

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/09/2017	10:45
2	CAL2	1	Water			7.02	05/09/2017	10:46
3	CAL3	1	Water			10.03	05/09/2017	10:47
4	ICV	1	Water			7.00	05/09/2017	10:49
5	CCV1	1	Water			2.02	05/09/2017	10:50
6	I3046-01	1	Water			6.08	05/09/2017	10:51
7	I3046-01DUP	1	Water			6.09	05/09/2017	10:52
8	I3046-02	1	Water			6.91	05/09/2017	10:53
9	CCV2	1	Water			12.02	05/09/2017	10:55

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3046

WorkList ID : 98258

Date : 5/9/2017 10:32:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3046-01 -F	pH	Cool 4 deg C	LIRO01	F62	TWP-13	05/08/2017	SM 4500-PH B
	Water	I3046-02 -F	pH	Cool 4 deg C	LIRO01	F62	TWP-11	05/08/2017	SM 4500-PH B

Date/Time 05/09/17 10:32 AM
 Received by: JP
 Relinquished by: JP

Date/Time 05/09/17 11:00 AM
 Received by: JP
 Relinquished by: JP