



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87328

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/09/2017	10:45
2	CAL2	1	Water			7.02	05/09/2017	10:46
3	CAL3	1	Water			10.03	05/09/2017	10:47
4	ICV	1	Water			7.00	05/09/2017	10:49
5	CCV1	1	Water			2.02	05/09/2017	11:00
6	I3046-09	1	Solid	20.03	20	7.32	05/09/2017	11:01
7	I3046-09DUP	1	Solid	20.05	20	7.33	05/09/2017	11:03
8	I3046-10	1	Solid	20.02	20	8.36	05/09/2017	11:05
9	I3046-11	1	Solid	20.05	20	7.84	05/09/2017	11:06
10	CCV2	1	Water			12.02	05/09/2017	11:09

LD 87328

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3046 WorkList ID : 98261 Date : 5/9/2017 10:34:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3046-09	Corrosivity	Cool 4 deg C	LIR001	F62	SB-11-COMP	05/08/2017	9045C
	Solid	I3046-10	Corrosivity	Cool 4 deg C	LIR001	F62	SB-12-COMP	05/08/2017	9045C
	Solid	I3046-11	Corrosivity	Cool 4 deg C	LIR001	F62	SB-13-COMP	05/08/2017	9045C

Date/Time 05/09/17
 Received by: SR
 Relinquished by: [Signature]

Date/Time 05/09/17- 5:00 PM
 Received by: CR
 Relinquished by: JO