

7471A


LB 87335

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44673



A= 0.0000e+000

B= 2.2427e-004

C= 2.3183e-003

Rho= 0.9999827

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	6	0.000	6				
0.2	0.200	0.206	0.006	910	0.0 %	910				
2.5	2.500	2.518	0.018	11219	0.0 %	11219				
5.0	5.000	4.956	-0.044	22087	0.0 %	22087				
7.5	7.500	7.495	-0.005	33409	0.0 %	33409				
10.0	10.000	10.021	0.021	44673	0.0 %	44673				

LB87335 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 09 May 2017 14:51:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	7471A	PPB	09 May 2017 14:57:47	S	Std
0.2 - 1	50.2	910	-	0.2000	7471A	PPB	09 May 2017 15:00:04	S	Std
2.5 - 1	52.5	11219	-	2.5000	7471A	PPB	09 May 2017 15:02:22	S	Std
5.0 - 1	55	22087	-	5.0000	7471A	PPB	09 May 2017 15:04:39	S	Std
7.5 - 1	57.5	33409	-	7.5000	7471A	PPB	09 May 2017 15:06:57	S	Std
10.0 - 1	510	44673	-	10.0000	7471A	PPB	09 May 2017 15:12:06	S	Std
ICV49 - 1	ICV49	17378	3.8996	-	7471A	PPB	09 May 2017 15:14:45	U	SMPL
ICB49 - 1	ICB49	-63	-0.0118	-	7471A	PPB	09 May 2017 15:17:01	U	SMPL
CCV68 - 1	CCV68	22312	5.0062	-	7471A	PPB	09 May 2017 15:21:49	U	SMPL
CCB68 - 1	CCB68	-95	-0.0190	-	7471A	PPB	09 May 2017 15:24:04	U	SMPL
CRA - 1	CRA	893	0.2026	-	7471A	PPB	09 May 2017 15:26:20	U	SMPL
HighStd - 1	HighStd	45252	10.1509	-	7471A	PPB	09 May 2017 15:28:36	U	SMPL
ChkStd - 1	ChkStd	31049	6.9656	-	7471A	PPB	09 May 2017 15:30:52	U	SMPL
PB98808BL - 1	PBS	-116	-0.0237	-	7471A	PPB	09 May 2017 15:33:08	U	SMPL
PB98808BS - 1	LCSS	15456	3.4686	-	7471A	PPB	09 May 2017 15:35:25	U	SMPL
I2872-13 - 1	PT-MET-SOIL	861214	HIGH	-	7471A	PPB	09 May 2017 15:37:42	X	
I3037-05 - 1	HY-SB408A	8208	1.8431	-	7471A	PPB	09 May 2017 15:39:32	U	SMPL
I3037-06 - 1	HY-SB408B	16833	3.7774	-	7471A	PPB	09 May 2017 15:41:59	U	SMPL
I3037-07 - 1	HY-SB408C	10378	2.3298	-	7471A	PPB	09 May 2017 15:44:17	U	SMPL
I3037-08 - 1	HY-SB409A	1214	0.2746	-	7471A	PPB	09 May 2017 15:46:32	U	SMPL
CCV69 - 1	CCV69	21040	4.7209	-	7471A	PPB	09 May 2017 15:48:48	U	SMPL
CCB69 - 1	CCB69	-100	-0.0201	-	7471A	PPB	09 May 2017 15:51:04	U	SMPL
I3037-09 - 1	HY-SB409B	12591	2.8261	-	7471A	PPB	09 May 2017 15:53:20	U	SMPL
I3040-04 - 1	TP-4-5COMP	843	0.1914	-	7471A	PPB	09 May 2017 15:55:36	U	SMPL
I3048-01 - 1	RO-4923	662	0.1508	-	7471A	PPB	09 May 2017 15:57:53	U	SMPL
I3048-01DUP - 1	RO-4923DUP	732	0.1665	-	7471A	PPB	09 May 2017 16:02:49	U	SMPL
I3048-01MS - 1	RO-4923MS	15776	3.5404	-	7471A	PPB	09 May 2017 16:05:05	U	SMPL
I3048-01MSD - 1	RO-4923MSD	15894	3.5668	-	7471A	PPB	09 May 2017 16:07:22	U	SMPL
I3049-01 - 1	DRUM-M4935	120	0.0292	-	7471A	PPB	09 May 2017 16:09:40	U	SMPL
I3048-01LX5 - 1		119	0.0290	-	7471A	PPB	09 May 2017 16:21:35	U	SMPL
I3048-01A - 1		18196	4.0831	-	7471A	PPB	09 May 2017 16:23:51	U	SMPL
CCV70 - 1	CCV70	21405	4.8028	-	7471A	PPB	09 May 2017 16:26:07	U	SMPL
CCB70 - 1	CCB70	-54	-0.0098	-	7471A	PPB	09 May 2017 16:28:23	U	SMPL
I2872-13DLX10 - 1	PT-MET-SOIL	192138	43.0925	-	7471A	PPB	09 May 2017 16:30:40	U	SMPL
I2872-13DLX100 - 1	PT-MET-SOIL	20044	4.4975	-	7471A	PPB	09 May 2017 16:42:14	U	SMPL
CCV71 - 1	CCV71	22282	4.9994	-	7471A	PPB	09 May 2017 16:49:04	U	SMPL
CCB71 - 1	CCB71	-67	-0.0127	-	7471A	PPB	09 May 2017 16:51:20	U	SMPL