



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87345

Temp.: 23.6 °C

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/09/2017	17:25
2	CAL2	1	Water			7.01	05/09/2017	17:26
3	CAL3	1	Water			10.02	05/09/2017	17:27
4	ICV	1	Water			7.00	05/09/2017	17:30
5	CCV1	1	Water			2.01	05/09/2017	17:31
6	I3062-15	1	Water			6.48	05/09/2017	17:32
7	I3062-15DUP	1	Water			6.49	05/09/2017	17:33
8	CCV2	1	Water			12.01	05/09/2017	17:35

198345

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 13062 WorkList ID : 98280 Date : 5/9/2017 5:06:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3062-15	pH	Cool 4 deg C	LIR001	G52	TWP-02	05/09/2017	SM 4500-PH B

Date/Time 05/09/17 5:06 PM
Received by: J
Relinquished by: CR

Date/Time 05/09/17 6:20 PM
Received by: CR
Relinquished by: J