


ISM02.4

LB 89351
INSTRUMENT ID: CV1

Linear

 μ Abs.:

46949



A= 0.0000e+000

B= 2.1416e-004

C= 3.9988e-002

Rho= 0.9998587

Accept=Accepted

slope
 $y = mx + b$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.037	0.037	-15	0.000	-15				
0.2	0.200	0.237	0.037	922	0.0 %	922				
2.5	2.500	2.461	-0.039	11305	0.0 %	11305				
5.0	5.000	4.962	-0.038	22983	0.0 %	22983				
7.5	7.500	7.408	-0.092	34405	0.0 %	34405				
10.0	10.000	10.095	0.095	46949	0.0 %	46949				

0/00

-

19

-2

-1

-1

1

NB

5/10/12

LB87351

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-15	-		0 ISM02.4	PPB	5/10/2017 9:25	Std
	0.2 S0.2	922	-		0.2 ISM02.4	PPB	5/10/2017 9:27	Std
	2.5 S2.5	11305	-		2.5 ISM02.4	PPB	5/10/2017 9:30	Std
	5 S5	22983	-		5 ISM02.4	PPB	5/10/2017 9:32	Std
	7.5 S7.5	34405	-		7.5 ISM02.4	PPB	5/10/2017 9:34	Std
	10 S10	46949	-		10 ISM02.4	PPB	5/10/2017 9:36	Std
ICV	ICV	18526	4.0075 -		ISM02.4	PPB	5/10/2017 9:40	SMPL
ICB	ICB	-84	0.022 -		ISM02.4	PPB	5/10/2017 9:42	SMPL
CCV072	CCV072	23673	5.1098 -		ISM02.4	PPB	5/10/2017 9:44	SMPL
CCB072	CCB072	-85	0.0218 -		ISM02.4	PPB	5/10/2017 9:46	SMPL
PB98809BL	PBS004	-27	0.0342 -		ISM02.4	PPB	5/10/2017 9:51	SMPL
I3016-01	PTS-01	1615	0.3859 -		ISM02.4	PPB	5/10/2017 9:53	SMPL
I3016-01DUP	PTS-01DUP	1312	0.321 -		ISM02.4	PPB	5/10/2017 9:56	SMPL
I3016-01MS	PTS-01MS	12726	2.7654 -		ISM02.4	PPB	5/10/2017 9:58	SMPL
CCV073	CCV073	23568	5.0873 -		ISM02.4	PPB	5/10/2017 10:00	SMPL
CCB073	CCB073	-72	0.0246 -		ISM02.4	PPB	5/10/2017 10:03	SMPL