



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87353

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	05/10/2017	09:30
2	CAL2	1	Water			7.02	05/10/2017	09:31
3	CAL3	1	Water			10.03	05/10/2017	09:32
4	ICV	1	Water			6.99	05/10/2017	09:33
5	CCV1	1	Water			2.01	05/10/2017	09:34
6	I3059-03	1	Solid	20.02	20	8.55	05/10/2017	09:36
7	I3059-03DUP	1	Solid	20.04	20	8.56	05/10/2017	09:37
8	I3062-16	1	Solid	20.02	20	7.42	05/10/2017	09:38
9	I3062-17	1	Solid	20.01	20	7.05	05/10/2017	09:40
10	I3062-18	1	Solid	20.03	20	7.53	05/10/2017	09:41
11	I3062-19	1	Solid	20.01	20	7.32	05/10/2017	09:42
12	I3062-20	1	Solid	20.05	20	7.55	05/10/2017	09:43
13	I3062-21	1	Solid	20.03	20	8.00	05/10/2017	09:45
14	I3062-22	1	Solid	20.01	20	7.62	05/10/2017	09:46
15	CCV2	1	Water			12.01	05/10/2017	09:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3062

WorkList ID : 98304

Date : 5/10/2017 9:04:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2017	Solid	I3059-03	Corrosivity	Cool 4 deg C	PSEG01	F51	JC-SWITCH-DRUM-SPILL(CC	05/09/2017	9045C
	Solid	I3062-16	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-10-COMP	05/08/2017	9045C
	Solid	I3062-17	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-09-COMP	05/08/2017	9045C
	Solid	I3062-18	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-08-COMP	05/08/2017	9045C
	Solid	I3062-19	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-14-COMP	05/09/2017	9045C
	Solid	I3062-20	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-06-COMP	05/08/2017	9045C
	Solid	I3062-21	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-02-COMP	05/09/2017	9045C
	Solid	I3062-22	Corrosivity	Cool 4 deg C	LIRO01	G52	SB-03-COMP	05/09/2017	9045C

Date/Time 05-10-17 01:04 AM
 Received by: JD
 Relinquished by: JD

Date/Time 05-10-17 12:30 PM
 Received by: CP
 Relinquished by: JD