

LB 87355

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/10/2017 9:58

Test: Phenolics-

wfll

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.994	0.0	0.111	
ICB1	0.002	0.0	0.002	
CCV1	1.001	0.0	0.112	
CCB1	-0.001	0.0	0.002	
PB98815BL	-0.000	0.0	0.002	
PB98815BS	1.003	0.0	0.112	
I3046-01	0.001	0.0	0.002	
I3046-01DUP	-0.002	0.0	0.002	
I3046-01MS	0.949	0.0	0.106	
I3046-01MSD	0.958	0.0	0.107	
I3046-02	-0.003	0.0	0.002	
CCV2	0.998	0.0	0.111	
CCB2	-0.000	0.0	0.002	

N	13
Mean	0.454
SD	0.5110
CV%	112.62

CB 87355

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed May 10 09:22:26 2017

Wed May 10 09:57:21 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0514	mg/l	5/10/2017 9:22:26	
0.1PPM	A	Phenolics-	P	0.1015	mg/l	5/10/2017 9:22:27	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	5/10/2017 9:22:28	
0.50PPM	A	Phenolics-	P	0.4995	mg/l	5/10/2017 9:22:29	
1.0PPM	A	Phenolics-	P	0.9919	mg/l	5/10/2017 9:22:30	
2.0PPM	A	Phenolics-	P	2.0038	mg/l	5/10/2017 9:22:31	
ICV1	S	Phenolics-	P	0.9937	mg/l	5/10/2017 9:53:22	
ICB1	S	Phenolics-	P	0.0015	mg/l	5/10/2017 9:53:23	
CCV1	S	Phenolics-	P	1.0014	mg/l	5/10/2017 9:53:24	
CCB1	S	Phenolics-	P	-0.0011	mg/l	5/10/2017 9:53:25	
PB98815BL	S	Phenolics-	P	-0.0002	mg/l	5/10/2017 9:53:26	
PB98815BS	S	Phenolics-	P	1.0029	mg/l	5/10/2017 9:53:27	
I3046-01	S	Phenolics-	P	0.0007	mg/l	5/10/2017 9:53:28	
I3046-01DUP	S	Phenolics-	P	-0.0021	mg/l	5/10/2017 9:53:29	
I3046-01MS	S	Phenolics-	P	0.9485	mg/l	5/10/2017 9:53:30	
I3046-01MSD	S	Phenolics-	P	0.9583	mg/l	5/10/2017 9:53:31	
I3046-02	S	Phenolics-	P	-0.0026	mg/l	5/10/2017 9:53:32	
CCV2	S	Phenolics-	P	0.9975	mg/l	5/10/2017 9:53:33	
CCB2	S	Phenolics-	P	-0.0004	mg/l	5/10/2017 9:57:20	

LB 87355

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

5/10/2017 9:25

Test Phenolics-


Accepted 5/10/2017 9:25

Factor 9.134

Bias 0.002

Coeff. of det. 0.999969

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0514	0.0500	
2	Phenol-2	0.013	0.1015	0.1000	
3	Phenol-2	0.030	0.2519	0.2500	
4	Phenol-2	0.057	0.4995	0.5000	
5	Phenol-2	0.111	0.9919	1.0000	
6	Phenol-2	0.221	2.0038	2.0000	