


ISM02.4

LB 87356
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43251



A= 0.0000e+000

B= 2.3161e-004

C= 3.5636e-002

Rho= 0.9999072

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.028	0.028	-31	0.000	-31					1
0.2	0.200	0.209	0.009	749	0.0 %	749					4
2.5	2.500	2.525	0.025	10749	0.0 %	10749					1
5.0	5.000	4.897	-0.103	20988	0.0 %	20988					-2
7.5	7.500	7.488	-0.012	32176	0.0 %	32176					0
10.0	10.000	10.053	0.053	43251	0.0 %	43251					1

103
5/10/12

LB87356

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-31	-		0 ISM02.4	PPB	5/10/2017 10:39	Std
	0.2 S0.2	749	-		0.2 ISM02.4	PPB	5/10/2017 10:41	Std
	2.5 S2.5	10749	-		2.5 ISM02.4	PPB	5/10/2017 10:43	Std
	5 S5	20988	-		5 ISM02.4	PPB	5/10/2017 10:46	Std
	7.5 S7.5	32176	-		7.5 ISM02.4	PPB	5/10/2017 10:48	Std
	10 S10	43251	-		10 ISM02.4	PPB	5/10/2017 10:50	Std
ICV	ICV	16493	3.8555 -		ISM02.4	PPB	5/10/2017 10:53	SMPL
ICB	ICB	-120	0.0078 -		ISM02.4	PPB	5/10/2017 10:55	SMPL
CCV074	CCV074	20854	4.8656 -		ISM02.4	PPB	5/10/2017 10:58	SMPL
CCB074	CCB074	-117	0.0085 -		ISM02.4	PPB	5/10/2017 11:00	SMPL
PB98811BL	PBW003	46	0.0463 -		ISM02.4	PPB	5/10/2017 11:02	SMPL
I2844-09	MBDCA5	57	0.0488 -		ISM02.4	PPB	5/10/2017 11:05	SMPL
CCV075	CCV075	20947	4.8871 -		ISM02.4	PPB	5/10/2017 11:07	SMPL
CCB075	CCB075	-103	0.0118 -		ISM02.4	PPB	5/10/2017 11:09	SMPL