


7470A

LB87319
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36900



A= 0.0000e+000

B= 2.7134e-004

C= -6.5059e-002

Rho= 0.9999176

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	108	0.000	108				
0.2	0.200	0.165	-0.035	848	0.0 %	848				
2.5	2.500	2.541	0.041	9603	0.0 %	9603				
5.0	5.000	5.081	0.081	18965	0.0 %	18965				
7.5	7.500	7.502	0.002	27888	0.0 %	27888				
10.0	10.000	9.947	-0.053	36900	0.0 %	36900				

LB87359 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 May 2017 12:02:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	108	-	0.00007470A	PPB		10 May 2017 12:05:50	S	Std
0.2 - 1	50.2	848	-	0.20007470A	PPB		10 May 2017 12:07:59	S	Std
2.5 - 1	52.5	9603	-	2.50007470A	PPB		10 May 2017 12:10:07	S	Std
5.0 - 1	55	18965	-	5.00007470A	PPB		10 May 2017 12:12:15	S	Std
7.5 - 1	57.5	27888	-	7.50007470A	PPB		10 May 2017 12:14:23	S	Std
10.0 - 1	510	36900	-	10.00007470A	PPB		10 May 2017 12:16:30	S	Std
ICV50 - 1	ICV50	13955	3.7215	-7470A	PPB		10 May 2017 12:19:04	U	SMPL
ICB50 - 1	ICB50	8	-0.0629	-7470A	PPB		10 May 2017 12:21:10	U	SMPL
CCV76 - 1	CCV76	18496	4.9536	-7470A	PPB		10 May 2017 12:23:17	U	SMPL
CCB76 - 1	CCB76	185	-0.0149	-7470A	PPB		10 May 2017 12:25:23	U	SMPL
CRA - 1	CRA	900	0.1791	-7470A	PPB		10 May 2017 12:27:30	U	SMPL
HighStd - 1	HighStd	35948	9.6890	-7470A	PPB		10 May 2017 12:29:37	U	SMPL
ChkStd - 1	ChkStd	25320	6.8052	-7470A	PPB		10 May 2017 12:31:44	U	SMPL
PB98822BL - 1	PBW	118	-0.0330	-7470A	PPB		10 May 2017 12:33:52	U	SMPL
PB98822BS - 1	LCSW	13968	3.7250	-7470A	PPB		10 May 2017 12:35:59	U	SMPL
PB98822TB - 1	PB98787TB	188	-0.0140	-7470A	PPB		10 May 2017 12:38:07	U	SMPL
I3040-05 - 1	TP-4-5-COMP	199	-0.0111	-7470A	PPB		10 May 2017 12:40:14	U	SMPL
I3040-05DUP - 1	TP-4-5-COMPDUP	161	-0.0214	-7470A	PPB		10 May 2017 12:42:23	U	SMPL
I3040-05MS - 1	TP-4-5-COMPMS	13463	3.5880	-7470A	PPB		10 May 2017 12:44:32	U	SMPL
I3040-05MSD - 1	TP-4-5-COMPMSD	13286	3.5399	-7470A	PPB		10 May 2017 12:46:38	U	SMPL
CCV77 - 1	CCV77	18509	4.9571	-7470A	PPB		10 May 2017 12:48:45	U	SMPL
CCB77 - 1	CCB77	44	-0.0531	-7470A	PPB		10 May 2017 12:50:52	U	SMPL
I3048-04 - 1	RO-4923	194	-0.0124	-7470A	PPB		10 May 2017 12:52:59	U	SMPL
I3049-04 - 1	DRUM-M4935	145	-0.0257	-7470A	PPB		10 May 2017 12:55:06	U	SMPL
PB98823BL - 1	PBW	302	0.0169	-7470A	PPB		10 May 2017 12:57:13	U	SMPL
PB98823BS - 1	LCSW	13829	3.6873	-7470A	PPB		10 May 2017 12:59:20	U	SMPL
I3037-10 - 1	HY-SB409-GW	6196	1.6161	-7470A	PPB		10 May 2017 13:01:27	U	SMPL
I3037-10DUP - 1	HY-SB409-GWDUP	5200	1.3459	-7470A	PPB		10 May 2017 13:03:35	U	SMPL
I3037-10MS - 1	HY-SB409-GWMS	18418	4.9324	-7470A	PPB		10 May 2017 13:05:43	U	SMPL
I3037-10MSD - 1	HY-SB409-GWMSD	17647	4.7232	-7470A	PPB		10 May 2017 13:07:51	U	SMPL
I3039-04 - 1	INFLUENT-5-5-17	3	-0.0642	-7470A	PPB		10 May 2017 13:10:00	U	SMPL
CCV78 - 1	CCV78	18801	5.0364	-7470A	PPB		10 May 2017 13:12:07	U	SMPL
CCB78 - 1	CCB78	41	-0.0539	-7470A	PPB		10 May 2017 13:14:14	U	SMPL
I3040-05LX5 - 1		121	-0.0322	-7470A	PPB		10 May 2017 13:18:23	U	SMPL
I3040-05A - 1		14352	3.8292	-7470A	PPB		10 May 2017 13:20:30	U	SMPL
I3037-10LX5 - 1		1412	0.3181	-7470A	PPB		10 May 2017 13:22:36	U	SMPL
I3037-10A - 1		19127	5.1248	-7470A	PPB		10 May 2017 13:24:43	U	SMPL
CCV79 - 1	CCV79	18629	4.9897	-7470A	PPB		10 May 2017 13:26:50	U	SMPL
CCB79 - 1	CCB79	195	-0.0121	-7470A	PPB		10 May 2017 13:28:58	U	SMPL