

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87363

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54805
calibration std. phosphate 0.5 ppm	WP54804
calibration std. phosphate 0.3 ppm	WP54803
calibration std. 0 ppm	WP54800
phosphate ICV std.	WP54807
calibration std. phosphate 0.1 ppm	WP54802
calibration std. phosphate 0.05 ppm	WP54801
Combined reagent	WP54809
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate CCV std.	WP54806
Ascorbic Acid	WP54808

Intercept: -0.0007Slope: 0.6494Regression: 0.999779

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/10/2017	11:13
2	CAL2	0.05	1	50	50	0.032	05/10/2017	11:13
3	CAL3	0.10	1	50	50	0.065	05/10/2017	11:14
4	CAL4	0.30	1	50	50	0.186	05/10/2017	11:14
5	CAL5	0.50	1	50	50	0.332	05/10/2017	11:15
6	CAL6	1.00	1	50	50	0.647	05/10/2017	11:15



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87363

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	05/10/2017	11:16
2	ICB		1	50	50	0.000	0.001	05/10/2017	11:16
3	CCV1	0.50	1	50	50	0.325	0.502	05/10/2017	11:17
4	CCB1		1	50	50	0.000	0.001	05/10/2017	11:17
5	PB98848BL		1	50	50	0.000	0.001	05/10/2017	11:17
6	PB98848BS	0.50	1	50	50	0.323	0.498	05/10/2017	11:18
7	I3026-01		1	50	50	1.554	2.394	05/10/2017	11:18
8	I3026-01DUP		1	50	50	1.557	2.399	05/10/2017	11:19
9	I3026-01MS	0.50	1	50	50	1.686	2.597	05/10/2017	11:19
10	I3026-01MSD	0.50	1	50	50	1.680	2.588	05/10/2017	11:20
11	I3026-01		50	50	50	0.195	0.301	05/10/2017	11:20
12	I3026-01DUP		50	50	50	0.193	0.298	05/10/2017	11:21
13	I3026-01MS	0.50	50	50	50	0.201	0.311	05/10/2017	11:21
14	I3026-01MSD	0.50	50	50	50	0.200	0.309	05/10/2017	11:22
15	CCV2	0.50	1	50	50	0.329	0.508	05/10/2017	11:22
16	CCB2		1	50	50	0.000	0.001	05/10/2017	11:23