


245.1

LAB 7364
INSTRUMENT ID: CV2

Linear

 μ Abs.:

32938



A= 0.0000e+000

B= 3.0211e-004

C= -2.9996e-002

Rho= 0.9998965

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	73	0.000	73				
0.2	0.200	0.182	-0.018	702	0.0 %	702				
2.5	2.500	2.500	0.000	8375	0.0 %	8375				
5.0	5.000	5.003	0.003	16661	0.0 %	16661				
7.5	7.500	7.601	0.101	25260	0.0 %	25260				
10.0	10.000	9.921	-0.079	32938	0.0 %	32938				

LB87364 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 10 May 2017 14:25:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	73	-	0.0000	245.1	PPB	10 May 2017 14:31:18	S	Std
0.2 - 1	S0-2	702	-	0.2000	245.1	PPB	10 May 2017 14:33:26	S	Std
2.5 - 1	S2-5	8375	-	2.5000	245.1	PPB	10 May 2017 14:35:34	S	Std
5.0 - 1	S5	16661	-	5.0000	245.1	PPB	10 May 2017 14:37:43	S	Std
7.5 - 1	S7-5	25260	-	7.5000	245.1	PPB	10 May 2017 14:39:51	S	Std
10.0 - 1	S10	32938	-	10.0000	245.1	PPB	10 May 2017 14:41:58	S	Std
ICV51 - 1	ICV51	12825	3.8446	-245.1		PPB	10 May 2017 14:44:33	U	SMPL
ICB51 - 1	ICB51	5	-0.0285	-245.1		PPB	10 May 2017 14:46:39	U	SMPL
CCV80 - 1	CCV80	16544	4.9681	-245.1		PPB	10 May 2017 14:48:46	U	SMPL
CCB80 - 1	CCB80	10	-0.0270	-245.1		PPB	10 May 2017 14:50:52	U	SMPL
CRA - 1	CRA	686	0.1773	-245.1		PPB	10 May 2017 14:52:59	U	SMPL
HighStd - 1	HighStd	32765	9.8686	-245.1		PPB	10 May 2017 14:55:05	U	SMPL
ChkStd - 1	ChkStd	22876	6.8811	-245.1		PPB	10 May 2017 14:57:12	U	SMPL
PB98821BL - 1	PBW	-21	-0.0363	-245.1		PPB	10 May 2017 14:59:20	U	SMPL
PB98821BS - 1	LCSW	12835	3.8476	-245.1		PPB	10 May 2017 15:01:28	U	SMPL
I3046-01 - 1	TWP-13	164	0.0195	-245.1		PPB	10 May 2017 15:08:17	U	SMPL
I3046-01DUP - 1	TWP-13DUP	137	0.0114	-245.1		PPB	10 May 2017 15:10:25	U	SMPL
I3046-01MS - 1	TWP-13MS	13490	4.0455	-245.1		PPB	10 May 2017 15:12:33	U	SMPL
I3046-01MSD - 1	TWP-13MSD	14575	4.3733	-245.1		PPB	10 May 2017 15:14:42	U	SMPL
I3046-02 - 1	TWP-11	181	0.0247	-245.1		PPB	10 May 2017 15:16:49	U	SMPL
CCV81 - 1	CCV81	16704	5.0165	-245.1		PPB	10 May 2017 15:18:55	U	SMPL
CCB81 - 1	CCB81	22	-0.0234	-245.1		PPB	10 May 2017 15:21:02	U	SMPL
I3053-01 - 1	32220008-3	197	0.0295	-245.1		PPB	10 May 2017 15:23:09	U	SMPL
I3046-01LX5 - 1		79	-0.0061	-245.1		PPB	10 May 2017 15:25:16	U	SMPL
I3046-01A - 1		13602	4.0793	-245.1		PPB	10 May 2017 15:27:23	U	SMPL
CCV82 - 1	CCV82	16738	5.0267	-245.1		PPB	10 May 2017 15:29:30	U	SMPL
CCB82 - 1	CCB82	22	-0.0234	-245.1		PPB	10 May 2017 15:31:38	U	SMPL